




MINUTES OF THE TECHNICAL COMMITTEE MEETING HELD ON DT. 20.04.2024 AT 4:30 PM REGARDING CORRIGENDUM ON TECHNICAL SPECIFICATION BASED ON QUERIES REQUESTED BY PROSPECTIVE BIDDERS AGAINST THE AMENDMENT PUBLISHED VIDE NOTIFICATION NUMBER- 2553 DTD. 08.04.2024  
TENDER FLOATED FOR Supply, Installation & Commissioning of Pediatric Equipment on Rate Contract Basis Bid Reference No. OSMCL/2023-24/Pediatric Equipment/11

Notification no. 3099 OSMCL/2024,		Item Name : Baby Incubator	Date- 09.05.2024
Sl.	Amendment Request / Queries raised by the prospective bidders	Original Technical Specifications	Amendment On Technical Specification (NOTIFICATION NUMBER- 2553,Dtd. CORRIGENDUM ON AMENDMENT
1	The opening and closing of hood should be electrical and automatically operated, quick and simple	<b>Page NO-74 TECHNICAL SPECIFICATION: Point No-2:</b> The opening and closing of hood should be mechanical, quick and simple, also be possible when device is switched off.	<b>Amended as:</b> The opening and closing of hood should be <b>electrical and automatically operated OR mechanically operated</b> , quick and simple.
2	Weight of the device without accessories should be 150 Kg or less.	<b>Page NO-74 TECHNICAL SPECIFICATION: Point No-8:</b> Weight of the device without accessories should be 140 Kg or less.	<b>Amended as:</b> Weight of the device without accessories should be 150 Kg or less.
3	Radiant warmer should be with 1 high quality heating element for an optimal and even heat distribution on the mattress area, automatic adaption of radiant warmer when mattress is tilted	<b>Page NO-75 TECHNICAL SPECIFICATION: Point No-11:</b> Radiant warmer should be with 2 heating elements for an optimal and even heat distribution on the mattress area, automatic adaption of radiant warmer when mattress is tilted.	<b>Amended as:</b> Radiant warmer should be with <b>one or more</b> heating elements for an optimal and even heat distribution on the mattress area, automatic adaption of radiant warmer when mattress is tilted.
4	Should provide with a T Piece Infant Resuscitation System with inbuilt blender, Gas inlet hose pipe, Clamp for mounting and Patient tubing set -10 Nos.	<b>Page NO-75, TECHNICAL SPECIFICATION: Point No-18:</b> Should provide with a Infant Resuscitation System with Gas inlet hose pipe, Clamp for mounting and Patient tubing set -10 No.s.	<b>Amended as:</b> Should provide with a <b>T-Piece Infant Resuscitation System with inbuilt /External blender</b> , Gas inlet hose pipe, Clamp for mounting and Patient tubing set -10 Nos.
5	The system should have data exchange and interfaces serial port USB port, and RJ 45 port or should be HL7 compliant for ENICU connectivity through charting	<b>Page NO-75 TECHNICAL SPECIFICATION: Point No-21:</b> The system should have data exchange and interfaces serial port,USB port, Nurse call system and RJ 45 port for service purpose..	<b>Amended as:</b> The system should have data exchange and interfaces serial port,USB port, and RJ 45 port <b>OR HL7 compliant</b> for connectivity purpose.


  
20/4/24  
Prof. Dr. Sunil Agarwal  
Dept. Of Paediatrics  
SCB MC&H, Cuttack


  
20.04.2024  
Prof. Dr. Arati Behera  
Dept. of Paediatrics  
SJMC&H, Puri


  
Asso. Prof. Dr. Swarupa Panda  
Dept. of Paediatrics  
SCB MC&H, Cuttack

MINUTES OF THE TECHNICAL COMMITTEE MEETING HELD ON DT. 20.04.2024 AT 4:30 PM REGARDING CORRIGENDUM ON TECHNICAL SPECIFICATION BASED ON QUERIES REQUESTED BY PROSPECTIVE BIDDERS AGAINST THE AMENDMENT PUBLISHED VIDE NOTIFICATION NUMBER- 2553 DTD. 08.04.2024  
TENDER FLOATED FOR Supply, Installation & Commissioning of Pediatric Equipment on Rate Contract Basis Bid Reference No. OSMCL/2023-24/Pediatric Equipment/11

Notification no. 3098 OSMCL/2024,		Item Name : Radiant Warmer		Date- 04.05.2024
Sr.	Amendment Request / Queries raised by the prospective bidders	Original Technical Specifications	Amendment On Technical Specification (NOTIFICATION NUMBER- 2553, Dtd. 08.04.2024 )	CORRIGENDUM ON AMENDMENT
1	Request you to remove this point	Page NO-68 TECHNICAL SPECIFICATION: Point No-12: The height of the warmer should be adjustable for Different types of bed.		<b>Amended as:</b> The technical specification mentioned at point no 12, i.e The height of the warmer should be adjustable for Different types of bed is removed.
2	The Height of the warmer should be electronically adjustable.	Page NO-68 TECHNICAL SPECIFICATION: Point No-12: The height of the warmer should be adjustable for different types of bed.		

  
20/4/24  
Prof. Dr. Sunil Agarwal  
Dept. Of Paediatrics  
SCB MC&H, Cuttack

  
20.04.2024  
Prof. Dr. Arati Behera  
Dept. of Paediatrics  
SJMC&H, Puri

  
Asso. Prof. Dr. Swarupa Panda  
Dept. of Paediatrics  
SCB MC&H, Cuttack