



Tender Reference No.: OSMCL/2018-19/EQP-Haemoglobinometer/05

Tender Name: e-Tender for Supply of Haemoglobinometer

**LIST OF BIDDERS WHICH ARE TECHNICALLY NON RESPONSIVE AND NOT QUALIFIED FOR PRICE BID OPENING
(EQUIPMENT-WISE)**

Sl.	Name of Bidder	Reason of Rejection
1	ASPEN DIAGNOSTICS PRIVATE LIMITED	NOT QUALIFIED as : The accuracy of the haemoglobin result during demonstration is below the accuracy range as per tender specifications.
2	Axess Medicom Pvt Ltd	NOT QUALIFIED as: 1. The audited turnover of manufacturer M/s Wrig Nanosystem is not qualifying as per the pre-qualification criteria of the tender. 2. The accuracy of the haemoglobin result during demonstration is below the accuracy range as per tender specifications.
3	Medico International	NOT QUALIFIED as: The submitted proof of performance is not sufficient as per tender terms (i.e. proof of supply to any Govt. organization / Private Hospitals/ Government Corporations/ Corporate Hospitals / PSU Hospitals / UN Agencies in India)
4	ARKRAY Healthcare Pvt. Ltd.	NOT QUALIFIED as: 1. The accuracy of the haemoglobin result during demonstration is below the accuracy range as per tender specifications. 2. The submitted proof of performance is not sufficient as per tender terms (i.e. proof of supply to any Govt. organization / Private Hospitals/ Government Corporations/ Corporate Hospitals / PSU Hospitals / UN Agencies in India)

Sd/-
Managing Director
OSMCL