

CORRIGENDUM

The Corrigendum has been made to the earlier published Pre-bid amendment point no- 25, 27, 34 & 39 to the earlier published Pre-bid amendment for Provision of 1.5T MRI services at AHRCC-Cuttack & Capital Hospital-BBSR in PPP mode (**Bid Ref. No. - OSMCL/2018-19/EQP-MRIPPP/35**) which may be read as below,

25. Amended as **“Dedicated Knee Coil / Coil with dedicated positioners for high resolution knee/ extremity coil should be provided with the system”**.

27. Amended as **“Dedicated Shoulder Coil / Coil with dedicated positioners for high resolution shoulder or extremity imaging with at least 6 channels”**.

34. Amended as **“Single sequence with multi B value diffusion imaging”**

39. Amended as **“Noise (in dBA) Maximum 120 dBA”**

The rest portions of the earlier published Pre-bid amendments for tender for Provision of 1.5T MRI services at AHRCC-Cuttack & Capital Hospital-BBSR in PPP mode remains unchanged.

**Sd/-
Managing Director
OSMC**