



ODISHA STATE MEDICAL CORPORATION LTD.
(A Govt. of Odisha Undertaking)
(Assuring Quality, Saving Lives)

Phone No.: (0674) 2380660,
Website: www.osmcl.nic.in,
E-mail: logistics.osmcl@gmail.com,
logistics.osmcl.od@nic.in

Letter No. 13608 /OSMC/F. No. I-81/2019

Dt. 28/11/2019

PRE-BID CORRIGENDUM

FOR SUPPLY OF SURGICAL, SUTURE & OTHER ITEMS FOR THE F.Y- 2019-20

Bid Ref. No.:- OSMCL/2019-20/(SUT.-SURG. & OTHER)-DHS/06

Sl. No	Item Name/ Particulars	Details as mentioned in the tender document	Queries raised by the prospective bidders	Clarifications/ Amendments in response to the queries.
Technical Queries				
1	Section-IV Item Sl. No. 80 Column No. 4 Page No.- 25	S02118 - Disposable Delivery Kit (ETO Sterilised) 1. Carbolic Soap, 25gm - 1 piece (Lifebuoy, Dettol etc.) 2. Sterilised Gloves:- Two pairs (6 ½&7 size) pre-powdered, BIS specifications gloves, surgical rubber made of Hypoallergic latex. 100% electronically tested, sterilised by Gamma Radiation / ETO, ISI Marked, IS No. 13422-92 = 1 pair 3. Surgical Blade, pre sterilised with Gamma radiation: Size – 24, ISI	1. Self declaration (notarised) to be submitted by the firm instead of market standing by licensing authority. (5.2.9) 2. Self declaration (notarised) to be submitted by the firm for No Conviction Certificate instead of submission of the same from licensing authority.(5.2.10) 3. Extension of delivery period upto 120 days instead of 70	1. Clarification Refer bid clause No. 5.2.1 (c) & 5.2.4 2. No Change. 3. No Change



ODISHA STATE MEDICAL CORPORATION LTD.
(A Govt. of Odisha Undertaking)
(Assuring Quality, Saving Lives)

Phone No.: (0674) 2380660,
 Website: www.osmcl.nic.in,
 E-mail: logistics.osmcl@gmail.com,
logistics.osmcl.od@nic.in

	<p>Marked 3319:1995 = 1 No.</p> <p>4. Sterilised umbilical cord clamp:-PVC, Gamma sterilised = 1 No.</p> <p>5. Plastic Sheet Perineal 4x6 ft. & couch 4x6 ft.) - 1 piece each</p> <p>6. Sterilised Gauze:- 10"x10", IS No.759/1988 = 4 pieces</p> <p>7. Absorbent Cotton 100gm Net - 1 Roll (Packet)</p> <p>8. Mop Egyptian cotton (15x15 inch) - 2 Nos.</p> <p>9. Neonatal Covering Sheet with hood (Egyptian Cotton) – 1 No. (2ft x 2ft)</p> <p>10. Water proof Apron – 1 No. (W: 3ft x H: 5ft)</p> <p>11. Roller Bandage (L: 4 mt. x W: 6 cm, IS-863-1988, Ends (column)-150, Picks (Row)-85/10 Sq. c.m) - 1 Piece</p> <p>12. Self-Sticking sterilized Maternity Napkin – 2 Nos (Belt less type, Size XL).</p>	<p>days.</p> <p>4. Any three years of experience in the past to be considered instead of last three years. (clause 5.2.8)</p> <p>5. Item No. 4, 6, 7, 11, 13 are covered under Medical Devices for which WHO GMP is not issued by licensing authority. So QMS certificate may be considered instead of WHO GMP.</p> <p>6. Carbolic soap can be supplied of other brand except Lifebuoy & Dettol.</p> <p>7. Sterilised Gloves can be provided with CE certification instead of BIS.</p> <p>8. Colour and thickness (in micron) of perennial sheet may be</p>	<p>4. No Change</p> <p>5. Clarification Refer clause no. 5.2.1(c) 5.2.4</p> <p>6. No Change</p> <p>7. No Change</p> <p>8. No Change</p>
--	---	---	---

		<p>13. Baby Mucous Sucker – 1 No.</p> <p>14. Triple layer filtered face mask - 1 No.</p>	<p>provided to finalize the kit cost.</p> <p>9. Colour and thickness (in micron) of Water proof apron may be provided to finalize the kit cost.</p> <p>10. As the Kit includes both drug and non drug items, ETO sterilization by loan licensee must be asked for drug items at the time of supply.</p> <p>11. As no license is issued for DD kits, request made for supply of the same under MSME registration.</p>	<p>9. No Change</p> <p>10. No Change</p> <p>11. Clarification Refer clause no. 5.2.1(c) 5.2.4</p>
2	<p>Section-IV</p> <p>Sl. No. 03.</p> <p>Column No. 4</p> <p>Page No. 12</p>	<p>S01003 –</p> <p>Black Braided Silk Size:- 0, length: (76-90) cm, ½ Circle Round Bodied 30 mm</p> <p>Needle (Sterilized) U.S.P, with CE certification</p>	<p>Request for supply with below specification: Black Braided Silk Size:- 0, length: (75-90) cm, ½ Circle Round Bodied/ Taper point 30 mm</p> <p>Needle BLACK,</p>	<p>No Change</p>



ODISHA STATE MEDICAL CORPORATION LTD.
(A Govt. of Odisha Undertaking)
(Assuring Quality, Saving Lives)

Phone No.: (0674) 2380660,
 Website: www.osmcl.nic.in,
 E-mail: logistics.osmcl@gmail.com,
logistics.osmcl.od@nic.in

			USFDA, CE Certification	
3	Section-IV Sl. No. 09. Column No. 4 Page No. 13	S01017 - Polypropylene Size:- 0, length: 70/90cm, 1/2 Circle Round Bodied 30mm (needle) U.S.P, with CE certification.	Request for supply with below specification: Polypropylene 70-90 cm, BLUE 27-30mm ½ Circle Taper point (heavy) with USFDA, CE certification	No Change
4	Section-IV Sl. No. 13. Column No. 4 Page No. 13	S01068 – Black Braided Silk Size:- 2-0, length: (76- 90) cm, ½ Circle Round Bodied 20-30mm needle (Sterilized) U.S.P, with CE certification.	Request for supply with below specification: Black Braided Silk 2-0, 75-90cm, 30mm, ½ circle round bodied/taper point, Black with USFDA, CE certification	No Change
5	Section-IV Sl. No. 14. Column No. 4 Page No. 14	S01069 – Black Braided Silk Size:-3-0, length: 76cm, ½ Circle Round Bodied 20 mm needle (Sterilized) U.S.P, with CE certification	Request for supply with below specification: Black Braided Silk 3-0 BLK 75-90cm 20-30mm ½ Circle taper point Black with USFDA CE Certification	No Change
6	Section-IV Sl. No. 18.	S01075 – Polypropylene	Request for supply with below specification:	No Change

	Column No. 4 Page No. 14	Size:- 2-0, length: 70cm, ½ Circle Round Bodied 20-30mm, with needle U.S.P, with CE certification	Polypropylene 2-0 70-75cm, BLUE 20-30mm ½ circle round bodied/taper point with USFDA, CE certification	
7	Section-IV Sl. No. 21. Column No. 4 Page No. 14	S01079 – Polyglactin Size:- 2-0, length: 90cm, ½ Circle Round Bodied 26-30mm (needle) U.S.P. with CE certification	Request for supply with below specification: Coated Braided polyglactin 910/Lactomer 91 2-0, 72cm, violet 26-30mm ½ circle round bodied/ taper point with CE/USFDA certification	No Change
8	Section-IV Sl. No. 82. Column No. 4 Page No. 26	S02134 – Salt Testing Kit (20ml Plastic Screwed cap bottle (reagent) and Iodine test colour chart (As per Tender Specification)	1. Three years average annual turnover criteria of One Crore or 90 lakhs of minimum turnover to be considered instead of minimum one crore turnover criteria for last three years. 2. Two bottles (of 10ml each) of reagent is suggested over	1. No Change 2. Amended. Two bottles of reagent (10ml

			one 20ml pack in order to have greater stability	each) or one bottle of 20ml of reagent may be supplied.
9	Section-IV Sl. No. 75. Column No. 4 Page No. 24	S02029 - Disposable Apron (Long body, sterile, disposable, Not less than 25 GSM Material, Colour- Green/Blue.)	Material of disposable apron is not specified: whether plastic or non-woven.	No Change
10	Section-IV Sl. No. 81. Column No. 4 Page No. 26	S02057 - Safe Delivery kit (As per tender specification)	Flannel cloth is proposed instead of Turkish towel wrapper with clip (for wrapping of baby) as flannel cloth is much softer, smother and warmer than Turkish cloth with rash and itching free characters.	No Change
11	Section-IV Sl. No. 56. Column No. 4 Page No. 26	S02027 - Makintosh Sheet (Size 90cm wide, length 100cms, ISI Marked, I.S. No.4135-1974 Amendment No.1 April 2005, Double colour, water proof. (The mackintosh should be of Grade A (i) as per IS specification 4135:1974. It should have 55% rubber polymer content of Proofing.)	Proposed for supply of Hospital rubber sheeting as per IS 8164: 1976 without reinforcing fabric Bicoloured with 65% rubber content to curb fungal attack and rashes to the patients skin during long use and skin friendly.	No Change

12	Section-IV Sl. No. 69 Column No. 4 Page No. 23	S02154 – Surgical Blades, Sizes 10 (Pre - sterile with Gamma Radiation for handle no. 4 sizes 10, IS No. : 3319:1995 with CE certification will be preferred)	Size 10 will be fitted only with handle no. 3 instead of handle No. 4	Amended. Handle Size “4” may be read as Handle size “3”
13	Section-IV Sl. No. 70 Column No. 4 Page No. 23	S02389 – Surgical Blades, Sizes 12 (Pre - sterile with Gamma Radiation for handle no. 4 sizes 12, IS No. : 3319:1995 with CE certification will be preferred)	Size 12 will be fitted only with handle no. 3 instead of handle No. 4	Amended. Handle Size “4” may be read as Handle size “3”
14	Section-IV Sl. No. 71 Column No. 4 Page No. 23	S02156 – Surgical Blades, Sizes 15 ((Pre - sterile with Gamma Radiation for handle no. 4 sizes 15, IS No. : 3319:1995 with CE certification will be preferred)	Size 15 will be fitted only with handle no. 3 instead of handle No. 4	Amended. Handle Size “4” may be read as Handle size “3”
Queries relating to general terms and conditions				
1	Section – VI Clause No. 6.28.9 Page no. 55	The successful bidder along with the copy of the invoice shall submit the copy of the Standard Quality certificate of analysis from their own laboratory / NABL	1. Non drug items in tender like BP blades, and sutures, needles etc only need sterility	1. No Change.



ODISHA STATE MEDICAL CORPORATION LTD.
(A Govt. of Odisha Undertaking)
(Assuring Quality, Saving Lives)

Phone No.: (0674) 2380660,
 Website: www.osmcl.nic.in,
 E-mail: logistics.osmcl@gmail.com,
logistics.osmcl.od@nic.in

		<p>accredited Laboratory / Government approved Laboratory as applicable</p> <p>with necessary protocols for every batch of items supplied to the supply points at the time of supply. The successful bidder also has to submit all the copies of the test reports to the Quality Assurance Division of OSMCL for future reference within 7 days of delivery.</p>	<p>certificate for each individual lot. Analytical test reports not to be asked for non drug products during supply.</p> <p>2. Oxygen flow meter & thermometer needs to be supplied with Lot No. & Mfg. Dt. Only.</p>	<p>2. No Change.</p>
2	<p>Section – VI</p> <p>Clause No. 6.31.3</p> <p>Page no. 57</p>	<p>Quality testing and Handling charges: 1.5 % of the purchase order value shall be collected from the approved supplier as the quality testing charges. But the Supply of each drug /consumable must be in minimum batches. If more than 1.5% of purchase order value is spent towards quality testing due to more number of batches, the extra cost will be collected from the supplier. The balance amount if any remaining due to less batch and bulk supply out of 1.5% will not be returned to the supplier.</p>	<p>Medical cylinders require test certificate provided by Dept. of explosives, Govt. of India, for which testing fees of the said items may be waived out.</p>	<p>No Change</p>



ODISHA STATE MEDICAL CORPORATION LTD.
(A Govt. of Odisha Undertaking)
(Assuring Quality, Saving Lives)

Phone No.: (0674) 2380660,
Website: www.osmcl.nic.in,
E-mail: logistics.osmcl@gmail.com,
logistics.osmcl.od@nic.in

General Clarification on Price Bid

- a) The bidders have to quote the prices as per the format mentioned in Price BOQ in the e-tender portal. The prices (Rates) are to be quoted per unit/ Kit/ cylinder/ film/ piece/ slide/ pair/ roll/ sheet/ tube/ catheter/ bag/ test etc.(as the case may be) and not in unit pack, the quoted price includes packing, forwarding / transportation & Door Delivery and excludes GST.
- b) The amendments mentioned above are to be treated as amendments in the technical specification(s) and term(s) and condition(s) of the above tender references. All other technical specifications and terms conditions remain unchanged.
- c) Since any text in the price BOQ can't be changed in the e-tender portal, the amendments mentioned in the specifications above are to be treated as amendments in the price bid/ BOQ also.

SD/-
Managing Director
Odisha State Medical Corporation Ltd.