

**PRE-BID CLARIFICATIONS / AMENDMENTS IN RESPONSE TO THE QUERIES RAISED BY PROSPECTIVE BIDDERS IN THE PRE-BID MEETING HELD ON 29.01.2019, 11.00 A.M IN THE CONFERENCE HALL OF OSMCL FOR THE TENDER OF HAEMOGLOBINOMETER**

**Bid Ref. No.: OSMCL/2018-19/EQP-Haemoglobinometer/23**

**Brief note on the Tender:**

- Date & Time of release of bid -19.01.2019, 3 PM.
- Date & time of Pre-bid meeting - 29.01.2019, 11 AM.
- Date & time of Online bid submission Start Date -07.02.2019, 3 PM.
- Last Date & Time of Online bid submission - 23.02.2019, 5 PM.
- Date & time of online Technical bid opening - 27.02.2019, 11 AM.
- Date of opening of Price Bid -To be informed to the qualified bidders.

**The tender was published in:**

**English Daily**

1. The Hindu
2. The Business Standard

**Odia Daily**

3. The Prameya
4. The Pragativadi
5. E-tender portal
6. OSMCL Website

- Total No. of prospective bidders/ representatives present in the pre-bid meeting are twelve **(12)**.
- The queries raised by the prospective bidders on the above mentioned tender references were thread barely discussed.
- The recommendations on commercial terms & conditions, technical specifications and pre-bid amendments with justifications are as listed below and are unanimously accepted by the committee :

<b>Item Name/ Particulars</b>	<b>Queries raised by the prospective bidders</b>	<b>Clarifications/ Amendments in response to the queries.</b>
Hemoglobinometer	Accuracy may be reduced to 90% from 96%	No Change
	Training of staff (medical, paramedical, technicians) may be restricted to CHC level instead of delivery locations.	<b>Amended</b> The supplier has to conduct training at the CHC level to the competent user and same shall be recorded and signed by the competent authority. A complete user operational guide

	shall have to be supplied along with the machine printed in Odia, Hindi and English language.
Requested for Self checking point: Self checking facility should be there whenever the analyser is turned on and at regular intervals thereafter if Analyser remains switched on.	No Change
Requested to clarify that both the device and the test strips/microcuvettes should be individually capable of operating in ambient temperature range of 10to 40°C.	No Change

#### Queries on commercial terms & conditions

	Requested not to change the Proof of Annual Average turnover criteria for manufacturer/Importer of Rs. 5 Crores or more	No Change
	Requested to change the Proof of Annual average turnover of the manufacturer/ importer to 2 Crores or more from 5 Crores or more in the last three (3) financial years on the following grounds : 1. In the last tender of OSMC, the proof of supply criteria has been reduced to 20% of the required quantity. Proportionately, M/s Axess Medicom has requested turnover to be reduced to Rs. 2 Crs. in this tender. 2. During previous 1 yr. the manufacturer, Wrig Nanosystems has supplied 17000 Nos. of Haemoglobinometers to various Government Hospitals in UP & MP. The bidder has enclosed supporting PO copies in order to proof his production and operational capabilities. 3. The bidder has enclosed a Memorandum from Ministry of Finance stating relaxation of prior turnover and experience criteria in Public Procurement for Start-up industries.	No Change

**N.B:-**

- 1. The amendments mentioned above are to be treated as amendments in the technical specification(s) and term(s) and condition(s) of the above tender reference. All other technical specifications and terms conditions remain unchanged.**
- 2. Since any text in the price BOQ can't be changed in the e-tender portal, the amendments mentioned in the specifications above are to be treated as amendments in the price bid/ BOQ also.**

**The meeting ended with thanks to all members.**

Sd/-

**Managing Director  
Odisha State Medical Corporation**