

PRE-BID CLARIFICATIONS / AMENDMENTS IN RESPONSE TO THE QUERIES RAISED BY PROSPECTIVE BIDDERS IN THE PRE-BID MEETING HELD ON 01.05.2018, 11.00 A.M IN THE CONFERENCE HALL OF OSMCL FOR THE SUPPLY & INSTALLATION OF ANATOMY EQUIPMENTS
Bid Ref. No.: OSMCL/2018-19/EQP-ANATOMY/01

Brief note on the Tender:

- Date & Time of release of bid -21.04.2018, 3 PM.
- Date & time of Pre-bid meeting - 01.05.2018, 11 AM.
- Date & time of Online bid submission Start Date - 07.05.2018, 3 PM.
- Last Date & Time of Online bid submission - 23.05.2018, 5 PM.
- Date & time of online Technical bid opening - 26.05.2018, 11 AM.
- Date of opening of Price Bid -To be informed to the qualified bidders.

The tender was published in:

1. The Indian Express
2. The Dharitri
3. E-tender Portal of NIC
4. OSMCL Website

- Total No. of prospective bidders/ representatives present in the pre-bid meeting are seven **(07)**.
- The queries raised by the prospective bidders on the above mentioned tender references were thread barely discussed.
- The recommendations on technical specifications and pre-bid amendments with justifications are attached as **Annexure-A** and are unanimously accepted by the committee.

ANNEXURE-A

Queries relating to Technical Specification (Item wise)

Sl.	Item Name/ Particulars	Queries raised by the prospective bidders	Original Technical Specification	Clarifications/ Amendments in response to the queries.
1	Trinocular Fluorescence Microscope	Section VII Technical specifications (Pg No-47) Optics system of German & Japanese type may be incorporated.	Optical system: Optics should DIN/ASTM international type.	No Change
		Section VII Technical specifications (Pg No-47) Numerical aperture with minimum 6 (sextuple) position coded turret nose piece with 10X, 20X, 40X, 100X oil emersion objectives lens may be asked.	Objectives: kindly mention the numerical aperture with minimum 4 nosepiece 10X, 20X, 40 X,100X Oil emersion	No Change
		Section VII Technical specifications (Pg No-48) Minimum 6 position filter wheel with fluorescence intensity may be asked.	Minimum 4 to 5 filter position or slot; with neutral density filter to control the fluorescence intensity	No Change

		Extra points needs to be added for better configuration Additional feature like upgradeable to 130W mercury fluorescence & DIC attachment can be incorporated.	No Change
2	Penta Head Teaching Microscope	Section VII Technical specifications (Pg No-46) Optics system of German & Japanese type may be incorporated.	Optical system: Optics should DIN/ASTM international type. No Change
		Section VII Technical specifications (Pg No-46) Illuminator may be LED instead of halogen bulb.	Built-in-Koehler illumination for transmitted light 12V, 30W halogen bulb (pre-centered) light Amended Illuminator should be Built-in-Koehler illumination for transmitted light 12V, 30W halogen bulb (pre-centered) light or LED.
		Section VII Technical specifications (Pg No-46) LBD & ND filter may not come as inbuilt.	Built-in filters (LBD-IF, ND6, ND25) Amended LBD & ND filter should be present with the microscope, either as inbuilt or separate as per the microscope design of the manufacturer.
		Section VII Technical specifications (Pg No-46) Digital 16MP CMOS sensor may be added along with CCD sensor.	Camera and Analysis Software: • Image Sensor: At least 16 Megapixel CCD Camera. • HD LCD display should be provided No Change
		Extra points needs to be added for better configuration Additional feature like upgradeable to 130W mercury fluorescence & DIC attachment can be incorporated.	No Change
3	Virtual Anatomy Dissection Table	Section VII Technical specifications (Pg No-59) The touch screen dimension may be 55 inch or better instead of 85inch. Or The touch screen dimension may be minimum 80inch or above. Or The tender specification should not be changed, it should be retained.	Screen Dimension: 85 inch(22.5"x81.5") or 216cm(57cmx207cm) Amended The touch screen display dimension should be minimum 80 inch or above.
		Extra points needs to be added for better configuration Product certification like ISO 13485 and CE & USFDA may be asked.	Amended The manufacturer should have ISO 13485 certification and the model should be CE and USFDA approved.
		Section VII Technical specifications (Pg No-59) Product dimension may be deleted.	Product Dimension : Length :2.2m Width:0.7m Height:0.9m Amended The product dimension should be as per the screen size of the quoted model.
		Extra points needs to be added for better configuration The dissection table may be easy to transport in corridor and lift.	No Change
		Extra points needs to be added for better configuration Imaging from PACS may be incorporated.	No Change
		Extra points needs to be added for better configuration May have electrical motorised adjustment of height and tilting from 180 to 90 degree for achieving black board positions.	Amended The feature of electrical motorised adjustment of height and tilting from 180 to 90 degree is also acceptable and allowed, but should not be a mandatory feature.

Queries and necessary recommendations relating to Techno-Commercial terms

Sl.	Item Name/ Particulars	Queries raised by the prospective bidders	Original Technical Specification	Clarifications/ Amendments in response to the queries.
1	All items of tender	6.7 Earnest Money Deposit (EMD): (pg no-20) The NSIC and MSME registered firms may be exempted from bid document and EMD fees as only Local MSE registered in Odisha has allowed for such exemption.	6.7.3 Local MSEs only registered in Odisha with the respective DICs, Khadi, Village, Cottage & Handicraft Industries, OSIC, NSIC shall be exempted from submission of EMD, subject to submission of the valid registration certificate from the concerned authority.	No Change
2	Anatomy models, Charts, Photographs	5.2 Pre qualification of Bidders: 5.2.1 (Vii) (pg no-14) The annual average turnover criteria may be reduced to average Rs.1 Crore instead of Rs.4Crore. for the items.	5.2.1 Manufacturer / Importers are eligible to participate in the bid provided, they fulfill the following conditions: (Vii) Proof of average annual turnover (Manufacturers/ Importer) of Rs. 4 Crores or more in the last 3 financial years i.e. 2014-15, 2015-16 and 2016-17 and it is certified by the Chartered Accountant as per the format at Format T8. 5.2.2 Authorized Distributors are eligible to participate in the bid provided: (ii) They should have Proof of Average annual turnover of Rs. 2 Crore or more in last 3 financial years i.e. 2015-16, 2016-17 and 2017-18 as per Format T8. In addition to this, the distributor shall also submit the average annual turnover of the manufacturer/importer of the item(s) as mentioned in 5.2.1 (vii) above.	<ul style="list-style-type: none"> The annual average turnover criteria of the manufacturer/importer for the items <u>Trinocular Fluorescence Microscope, Models, Charts & Photographs, Embalming Machine and Dissecting Microscope</u> is reduced to Rs.1 Crore or more in last three financial years. The annual average turnover criteria of the authorized distributor for the above mentioned items is reduced to Rs.50 Lakhs or more in last three financial years.
3.	Microtomes and Microscopes of the tender list of items	5.2 Pre qualification of Bidders: 5.2.1 (Vii) (pg no-14) The turnover may be reduced to Rs.2.5 Crore as 99% of the Indian manufacturers are coming under MSME category for these items.	5.2.1 Manufacturer / Importers are eligible to participate in the bid provided, they fulfill the following conditions: (Vii) Proof of average annual turnover (Manufacturers/ Importer) of Rs. 4 Crores or more in the last 3 financial years i.e. 2014-15,	<ul style="list-style-type: none"> The annual average turnover criteria of the manufacturer/importer for the items <u>Two body Mortuary Cooler, Semi automatic Rotary</u>

	Original Human Skeleton (Articulated & Disarticulated)	<p>5.2 Pre qualification of Bidders: 5.2.1 (vii) (pg no-14)</p> <p>The annual average turnover criteria may be reduced to Rs.3 Crore instead of Rs.4 Crore.</p>	<p>2015-16 and 2016-17 and it is certified by the Chartered Accountant as per the format at Format T8.</p> <p>5.2.2 Authorized Distributors are eligible to participate in the bid provided:</p> <p>(ii) They should have Proof of Average annual turnover of Rs. 2 Crore or more in last 3 financial years i.e. 2015-16, 2016-17 and 2017-18 as per Format T8. In addition to this, the distributor shall also submit the average annual turnover of the manufacturer/importer of the item(s) as mentioned in 5.2.1 (vii) above.</p>	<p>Microtome, Manual Rotary Microtome, Penta Head Teaching Microscope, Original Human Skeleton (Articulated) & Disarticulated), Virtual Anatomy Dissection Table and Connect Lab & Digital System is reduced to Rs.2 Crore or more in last three financial years.</p> <p>• The annual average turnover criteria of the authorized distributor for the above mentioned items is reduced to Rs.1 Crore or more in last three financial years.</p>
4	Original Human Skeleton (Articulated & Disarticulated)	EMD amount may be reduced to 50% of the tender value as these items are not a regular purchase items and suppliers are limited	<p>SECTION III TENDER SCHEDULE 3.1. Bid Details Earnest Money Deposit 7.Original Human Skeleton (Articulated)- Rs 1,00,000/- 8.Original Human Skeleton (Disarticulated)-Rs 2,00,000/-</p>	No Change

Note: The comprehensive warranty shall be **3 years** from the date of installation & commissioning for all the tender items **except for three items** i.e. 1)Charts, Models & Photographs, 2)Original Human Skeleton (Articulated) and 3)Original Human Skeleton (Dis-Articulated). For these 3 items, warranty is not applicable.

N.B:

1. The amendments mentioned above are to be treated as amendments in the technical specification(s) and term(s) and condition(s) of the above tender reference. All other technical specifications and terms conditions remain unchanged.

2. Since any text in the price BOQ can't be changed in the e-tender portal, the amendments mentioned in the specifications above are to be treated as amendments in the price bid/ BOQ also.

**Sd/
Managing Director, OSMC**