

PRE-BID CLARIFICATIONS / AMENDMENTS IN RESPONSE TO THE QUERIES RAISED BY PROSPECTIVE BIDDERS IN THE PRE-BID MEETING HELD ON 22.05.2018, 11.00 A.M IN THE CONFERENCE HALL OF OSMCL FOR THE TENDER OF HAEMOGLOBINOMETER

Bid Ref. No.: OSMCL/2018-19/EQP-Haemometer/05

Brief note on the Tender:

- Date & Time of release of bid -15.05.2018, 3 PM.
- Date & time of Pre-bid meeting - 22.05.2018, 11 AM.
- Date & time of Online bid submission Start Date -29.05.2018, 3 PM.
- Last Date & Time of Online bid submission - 20.06.2018, 5 PM.
- Date & time of online Technical bid opening - 27.06.2018, 11 AM.
- Date of opening of Price Bid -To be informed to the qualified bidders.

The tender was published in:

English Daily

1. The Times of India

Odia Daily

2. The Sambad
3. E-tender portal
4. OSMCL Website

- Total No. of prospective bidders/ representatives present in the pre-bid meeting are fifteen **(15)**.
- The queries raised by the prospective bidders on the above mentioned tender references were thread barely discussed.
- The recommendations on commercial terms & conditions, technical specifications and pre-bid amendments with justifications are as listed below and are unanimously accepted by the committee :

Item Name/ Particulars	Queries raised by the prospective bidders	Clarifications/ Amendments in response to the queries.
Hemoglobin ometer	The model may be both USFDA and European CE approved.	No Change
	The micro cuvette method may also be allowed.	No Change Clarified that only strip based method is allowed.
	Measurement range may be from 0 to 25.6gm/dl instead of 4 to 20gm/dl. Or	Amended The measurement range should be from 5 to 20gm/dL.

Measurement range may be from 5 to 20gm/dl instead of 4 to 20gm/dl.	
Hemoglobin & Hematocrit results from one sample may be available after 25 to 60 seconds.	No Change
The storage memory may be deleted as practical readings are real time only or Should have facility to transfer the test results to computer or external device.	No Change
Operating condition: May be capable of operating continuously in 10 to 35degC instead of 0 to 50degC.	No Change Clarified that the operating range is 10 to 40 degC as per the technical specifications of the tender.
The in-house test reports on parameters and factory calibration may be accepted.	No Change Clarified that the OEM should conform on factory calibration as per ICSH (International Council for Standardization in Haematology), however certificate or test report from NABL accredited lab must be furnished on accuracy of the product.
Rechargeable battery may be included.	No Change Clarification: Any of the options (Rechargeable/Nonrechargeable) available with the system is allowed however the backup should support atleast 500 tests.
Queries on commercial terms & conditions	
Criteria for evaluation of Annual average turnover of the manufacturer may be reduced to 1 year instead of 3 years.	No Change
Proof of supply may be reduced to 20% or 25% from 50% of the required quantity (for the device) or add pharma companies, private agencies, private hospitals along with other mentioned customers.	Amended Proof of supply should be of 20% of the required quantity (for the device only), however the proof of supply to any Govt. organization / Private Hospitals/ Government Corporations/ Corporate Hospitals / PSU Hospitals / UN Agencies in India

		shall be accepted with proper purchase order executed by direct manufacturer/importer/through distributor in last three years.
	The purchase order quantity for the device & test strips may be clarified.	No Change Clarified that the purchase order for the device shall be issued for the full quantity requirement as per tender schedule of requirement, however the purchase order for the test strips shall be issued in phased manner as per the requirement.

N.B:-

1. The amendments mentioned above are to be treated as amendments in the technical specification(s) and term(s) and condition(s) of the above tender reference. All other technical specifications and terms conditions remain unchanged.
2. Since any text in the price BOQ can't be changed in the e-tender portal, the amendments mentioned in the specifications above are to be treated as amendments in the price bid/ BOQ also.
3. Bidders shall submit any query or observation or representation on uploaded technical bids within 7 days of opening of Technical Bid. Bidders can submit their queries or observations or representations for their quoted items through email: proc.osmcl.od@nic.in or at OSMCL office. The query or observation or representation submitted must be supported with the letter from the principal manufacturer for the respective items along with supporting documents justifying their claims. **No representations shall be entertained by the Tender Inviting Authority after 07 days of Technical bid opening.** The Tender Inviting Authority may at its discretion accept or reject the representation and the decision of the Tender Inviting Authority shall be final and shall be binding on all bidders participating in this bid.

Amendment: The last of submission has been extended and revised dates are as follows:

Amended Last date for online technical bid submission - 28.06.2018, 5PM

Amended Last date for online technical bid opening - 03.07.2018, 11AM

The meeting ended with thanks to all members.