

PRE-BID CLARIFICATIONS / AMENDMENTS IN RESPONSE TO THE QUERIES RAISED BY PROSPECTIVE BIDDERS IN THE PRE-BID MEETING HELD ON 15.01.2019, 11.00 A.M IN THE CONFERENCE HALL OF OSMCL FOR THE TENDER FOR SUPPLY & INSTALLATION OF DENTAL EQUIPMENTS

Bid Ref. No.: OSMCL/2018-19/EQP-DENTAL(Cat.I)/18 & OSMCL/2018-19/EQP-DENTAL(Cat.II)/19

Brief note on the Tender:

- Date & Time of release of bid -07.01.2019, 3 PM.
- Date & time of Pre-bid meeting - 15.01.2019, 11 AM.
- Date & time of Online bid submission Start Date - 25.01.2019, 3 PM.
- Last Date & Time of Online bid submission - 11.02.2019, 5 PM.
- Date & time of online Technical bid opening - 16.02.2019, 11 AM.
- Date of opening of Price Bid -To be informed to the qualified bidders.

The tender was published in:

English Daily

1. The Times of India
2. The Hindu

Odia Daily

3. The Sambad
4. The Pragativadi
5. E-tender portal
6. OSMCL Website

- Total No. of prospective bidders/ representatives present in the pre-bid meeting are 04. However, 02 bidders have submitted their queries in writing on a later date.
- The queries raised by the prospective bidders on the above mentioned tender references were thread barely discussed.
- The recommendations on commercial terms & conditions, technical specifications and pre-bid amendments with justifications are as listed below and are unanimously accepted by the committee :

Item Name/ Particulars	Queries raised by the prospective bidders	Clarifications/ Amendments in response to the queries.
Dental Chair without Compressor	The point 3 way syringe for Air, Water and Spray – 2 Nos. may be deleted	Amended The point 3 way syringe for Air, Water and Spray – 2 Nos. is deleted
	Zero Back Ache Stools (01 No.) may be provided instead of 02 Nos.	Amended Zero Back Ache Stools (01 No.) shall be provided with the dental chair

Extra Oral Radiography Machine with CR System, printer and computer along with medical grade monitor	X-Ray Generator- mAs range may be amended to 2 to 300 mAs instead of 2 to 200 mAs.	Amended mAs range – 2 to 200 mAs or more
	X-Ray Tube - Small focus - 1 mm (small) & large focus – 1 & 2 mm should be provided	No Amendment
	CR System: The speed of the scanner may be specified.	Amended CR system should have capacity to process more than 60 or more cassette/films per hour
	CR System: “Rigid” Plate or “double side scanning technology	“Rigid” Plate or “double side scanning technology is deleted
	CR System: Printer should have a minimum resolution of 508 dpi/ppi.	No Amendment
	CR System: Online UPS may not be provided for the x-ray machine.	Amended Online UPS having adequate capacity for the working of the CR system and printer should be provided.
Dental Instrument Set	Size & quantity of each consisting items may be specified.	Amended The required size, type, quantity of each item of the set are specified in below as Annexure-I .
Queries on commercial terms & conditions		
	Submission of one technical bid for all items under Cat-I & II	Clarified that Bidder may prepare one technical bid document mentioning the details of all items, however the same document has to be uploaded every time for all the items intended for bidding.
	Submission of tender fee of : 1. Rs. 5600/- for all items under Category-I 2. Rs. 5600/- for all items under Category-II	Clarified that 1.The bidder has to submit one time tender fee of Rs. 5600/- (against any or all items) under Cat-I. 2.The bidder has to submit one time tender

14	Scissors		Tissue cutting, stitch cutting	5 each
15	Needle holder			2
16	Bone chisel (mono bevel)			3
17	Scalpel handle	No 3		3 each
18	Sterilizer Drum	Medium		3
19	Plaster spatula(Straight & curved)			1 each
20	Rubber bowl	Small / large		1 each
21	Suction tips			30
22	Cheek retractor	Paediatric / Adult		1 each
23	Mouth props (Adult & Pedo) with chain			1 each
25	Patient drape			10
26	Glass dappen dish			2
27	Mortar & pestle			2
28	Burs associated for contrangle hand piece			50
29	Composite kit with etchant and bonding agent			2
30	Composite syringes			2
31	Composite finishing & polishing kit			2
32	Diamond burs for air rotor hand piece (assorted)			4
33	GP point 15-80 assorted set			4
34	H file set assorted 15-40, 45-80 (21mm)			4
35	K file set assorted 15-40, 45-80 (21mm) (25mm)			4
36	Mylar strips (8mm,100 strips pack)			10
37	Polishing brush & cup			2
38	Applicator tips for bonding agent		Disposable	100

N.B:-

- 1. The amendments mentioned above are to be treated as amendments in the technical specification(s) and term(s) and condition(s) of the above tender reference. All other technical specifications and terms conditions remain unchanged.**
- 2. Since any text in the price BOQ can't be changed in the e-tender portal, the amendments mentioned in the specifications above are to be treated as amendments in the price bid/ BOQ also.**

Amendment: The last of submission has been extended and revised dates are as follows:

Amended Last date for online technical bid submission - 18.02.2019, 5PM

Amended Last date for online technical bid opening - 22.02.2019, 11AM

The meeting ended with thanks to the chair.

Sd/-

**Managing Director
OSMCL**