



Letter No: 9002

Date: 09-07-2020

## **BIDDERS' SELECTION / REJECTION LIST**

Short Quotation Notification No. 6387 Dtd. 07-05-2020

**SHORT NOTICE TENDER FOR SUPPLY OF NON-CONTACT INFRARED THERMOMETER**

### **I. LIST OF BIDDERS & THEIR OFFERED EQUIPMENT WHICH ARE TECHNICALLY RESPONSIVE AND QUALIFIED FOR PRICE BID OPENING (ITEM-WISE)**

Sl.	Name of the vendor	Quoted Make	Quoted Model with 510 k No.	Status of Bidder with reason of rejection
1	VAIDYA V & I INFRASTRUCTURE PVT. LTD.	Visiomed Technology Co. Ltd.	LX-26E (K160816)	QUALIFIED
2	E.G. KANTAWALLA PVT. LTD.	Shenzhen Pango Electronic Company Ltd.	PG-IRT1602 (K182597)	QUALIFIED
3	SOURYA MEDICAL SYSTEM	Alicn Medical	AET R1B1 (K180207)	QUALIFIED
4	MAKS ENERGY SOLUTIONS	Shenzhen Pacom Medical Instruments Co. Ltd.	PC868 (K172889)	QUALIFIED
5	SAI VALSA EXPORTS GLOBAL TEXTILES	HeTaiDa Technology Co Ltd	HTD-8808C (K171888)	QUALIFIED
6	CGN RESEARCH LABS PVT. LTD.	Guangzhou Berrcom Medical Device Co. Ltd	MD H30 (K191570)	QUALIFIED
7	NEXELLECE CONSULTING SERVICES PVT LTD.	HeTaiDa Technology Co Ltd	8818A Series (K171888)	QUALIFIED
8	FABKON TECHNICAL SYSTEMS	Alicn Medical (Shenzhen) INC	BBLOVE AET-R1B1 (K180207)	QUALIFIED
9	J.K. MEDICAL SYSTEM	Shenzhen Jiacom Technology Co. Ltd.	FR850 (K172874)	QUALIFIED
10	QUICK PAY PVT. LTD.	Shenzhen Pacom Medical Instruments Co. Ltd.	PC868 (K172889)	QUALIFIED
11	PRAKRITI ENTERPRISERS	Visiomed Technology Co. Ltd.	LX-26E (K160816)	QUALIFIED
12	PRIMETIME ENTERTAINMENT & MEDIA PVT. LTD.	Shenzhen Pacom Medical Instruments Co. Ltd.	PC868 (K172889)	QUALIFIED
13	RADHA ENTERPRISES	Alicn Medical Shenzhen ,INC	AET-R1B1 (K180207)	QUALIFIED
14	AGAM PHARMACEUTICAL	Xiamen Ants Bro Technology Co.Ltd	HA-650 (K190984)	QUALIFIED
15	RABINDRA SURGICALS PVT. LTD.	Alicn Medical Shenzhen ,INC	AET-R1B1 (K180207)	QUALIFIED

**II. LIST OF BIDDERS & THEIR QUOTED MODELS WHICH ARE TECHNICALLY  
NOT QUALIFYING WITH REASON OF REJECTION**

Sl.	Name of the vendor	Quoted Make	Quoted Model	Status of Bidder with reason of rejection
1	RIKI SECURITY SOLUTIONS PVT. LTD.	Hunan Honggao electronic technology co.ltd	HG03	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of (i) Valid ISO 13485 certificate. (ii) Valid CE certificate from a notified body.
2	AIMIL LTD.	Berrcom	JXB-178	<b>NOT QUALIFIED</b> because the quoted model is NOT USFDA approved with 510K number.
3	CONTINENTAL TELESERVICES	CP.PLUS	YS-ET05	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of Valid CE certificate issued from a notified body.
4	VARIETY VYAPAAR PVT. LTD.	Guizhou Tekang Electronics Co.Ltd	T168	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of (i) Valid ISO 13485 certificate. (ii) Valid CE certificate issued from a notified body.
5	AMARAVATI ENTERPRISES	FLIR system india Pvt.Ltd	EXTECH IR267	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of Valid CE certificate issued from a notified body.
6	SHRI MAHALAXMI INDUSTRIES	Hangzhou Qingyuan Medical Equipment Technology	Qy-EWQ-01	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of valid CE certificate issued from a notified body.
7	INDIAN SURGICAL EQUIPMENT CO. PVT. LTD.	Contec Medical system co.Ltd	TP 500	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of valid CE certificate issued from a notified body.
8	SAI INFORMATIQUES	Time watch infocom Pvt.ltd.	R1D2	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of (i) Valid ISO 13485 certificate. (ii) Valid CE certificate issued from a notified body.
9	QUANTUM HI-TECH MERCHANDISING PVT. LTD.	Hunan Honggao electronic technology co.ltd	HG01	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of Valid ISO 13485 certificate.

Sl.	Name of the vendor	Quoted Make	Quoted Model	Status of Bidder with reason of rejection
10	JAI RADISSON (I) PVT. LTD.	Hunan Honggao electronic technology co.ltd	HG01	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of (i) Valid ISO 13485 certificate. (ii) Valid CE certificate issued from a notified body.
11	VARN PHARMACEUTICALS	Hunan Honggao electronic technology co.ltd	HG01	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of (i) Valid ISO 13485 certificate. (ii) Valid CE certificate issued from a notified body.
12	SISCO LATEX PVT. LTD.	Jiangxi aicare medical technology.ltd.	A66	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of Valid CE certificate issued from a notified body.
13	SHIVAM CREATION	Shenzhen hezhizhou technology co.ltd.	HT-668	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of (i) Valid ISO 13485 certificate. (ii) Valid CE certificate issued from a notified body.
14	GLOBAL QUEST CONSULTANTS.	Shenzhen zheng kang technology co.ltd	Trade name: JZIKI, (Model name : NA)	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of Valid CE certificate issued from a notified body.
15	AMTEL COMMUNICATION PVT. LTD.	Hunan JRT medical Devices Technology Co.Ltd.	JRT-018	<b>NOT QUALIFIED</b> due to Non-submission of (i) Valid CE certificate issued from a notified body. (ii) Valid ISO 13485 certificate.
16	SB TRADE IN PVT. LTD.	Raoping County Meiyuer Electronics Factory	YNA-8000	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of Valid CE certificate issued from a notified body.
17	MEDIWAYS CHEMIST & COSMETICS	Raoping County Meiyuer Electronics Factory	YNA-8000	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of Valid CE certificate issued from a notified body.

Sl.	Name of the vendor	Quoted Make	Quoted Model	Status of Bidder with reason of rejection
18	SHREE BALAJI UDYOG	Harbin Xiande Technology Development Co.Ltd	GP-300	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of (i) Valid ISO 13485 certificate. (ii) Valid CE certificate issued from a notified body.
19	HONEY ENTERPRISES	Henan Bingzun industrial.co.Ltd.	BZ-R6	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of (i) Valid ISO 13485 certificate. (ii) Valid CE certificate issued from a notified body.
20	SHIREEN ELECTRONICS PVT. LTD.	Zhengzhou AiQURA intelligent technology Co.Ltd	Vmade AD 801	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of (i) Valid ISO 13485 certificate. (ii) Valid CE certificate issued from a notified body.
21	EMBEE SOFTWARE PVT. LTD.	QQZM	GP-300	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of (i) Valid ISO 13485 certificate. (ii) Valid CE certificate issued from a notified body.
22	SHEO INDUSTRIES PVT. LTD.	Shenzhen yostand technology co.ltd.	YS-ET05	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of (i) Valid ISO 13485 certificate. (ii) Valid CE certificate issued from a notified body.
23	RAJCOMM COMMUNICATION SYSTEMS PVT. LTD.	Vantage	W-F168-L5	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of (i) Valid ISO 13485 certificate. (ii) Valid CE certificate issued from a notified body.
24	HALELU GLOBAL	Aiqura	AD801	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of (i) Valid ISO 13485 certificate. (ii) Valid CE certificate issued from a notified body.

Sl.	Name of the vendor	Quoted Make	Quoted Model	Status of Bidder with reason of rejection
25	VERTEL INFOTECH PVT. LTD.	Tahoco Electronic Technology(fujian) Co.Ltd	THK-TOP01	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of (i) Valid ISO 13485 certificate. (ii) Valid CE certificate issued from a notified body.
26	PATTANAIAK SCIENTIFIC SUPPLY SYNDICATE	Cole-Parmer India Pvt.Ltd	39642-02	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of (i) Valid ISO 13485 certificate. (ii) Valid CE certificate issued from a notified body.
27	MOBEL INDIA PVT. LTD.	Shenzhen super global electronics co.ltd	YHKY-2000	<b>NOT QUALIFIED</b> due to Non-submission of (i) Valid CE certificate issued from a notified body. (ii) Valid ISO 13485
28	CLASSIC TECHNOLOGIES & TRADE PVT. LTD.	HTC	HTC SCAN II	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of (i) Valid ISO 13485 certificate. (ii) Valid CE certificate issued from a notified body.
29	KARTTINE HOLDINGS PVT. LTD.	Bionics Scientific Technologies Pvt. Ltd.	TG 1804	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of valid CE certificate issued from a notified body.
30	AAR AAR LOGISTICS & SOURCING	Jo Pharma	JO-600	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of (i) Valid ISO 13485 certificate. (ii) Valid CE certificate issued from a notified body.
31	HIMACHAL PHARMACEUTICALS	Shenzhen Super Global Electronics Co. Ltd.	XHKY-2000	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of valid CE certificate issued from a notified body.
32	BHARDWAJ HOSPITALITY SERVICES			<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of (i) Valid ISO 13485 certificate. (ii) Valid CE certificate issued from a notified body.

Sl.	Name of the vendor	Quoted Make	Quoted Model	Status of Bidder with reason of rejection
33	STATION SATCOM (P) LTD.	Hunan Honggao Electronic Technology Co. Ltd.	DIKANG HG 01	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number.
34	VISAKHA HEALTHCARE (I) PVT. LTD.	Dongguan Xintai Instrument Co. Ltd.	HT-820D	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of (i) Valid ISO 13485 certificate. (ii) Valid CE certificate issued from a notified body.
35	V. LINK SYSTEMS PVT. LTD.	Hunan JRT Medical Devices Co. Ltd.	JRT-019	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of valid CE certificate issued from a notified body.
36	UMANG ASSOCIATES PVT. LTD.	Hunan Honggao Electronic Technology Co. Ltd.	(Multiple model quoted) HG 01, HG 02	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Multiple model quoted.
37	PROGILITY TECHNOLOGIES PVT. LTD.	Shandong Bittel Intelligent Technology Co. Ltd.	BITEWJ-02	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of (i) Valid ISO 13485 certificate. (ii) Valid CE certificate issued from a notified body.
38	VENTOX HEALTHCARE LTD.	Shenzhen Finicare Co. Ltd.	FC-IR 100	<b>NOT QUALIFIED</b> because the quoted model has deviation in the technical specification, i.e. as per the technical characteristics of the subject model in USFDA certificate, the measurement distance is 0 cm instead of 5 cm or better.
39	SANASH IMPEX PVT. LTD.	Anshun Intech Technology Co. Ltd.	YT-EWQ-04	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of (i) Valid ISO 13485 certificate. (ii) Valid CE certificate issued from a notified body.
40	GLADIOLUX INFRASTRUCTURES SERVICES	Analytical Technologies Ltd.	IR 3000 Series	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of valid CE certificate issued from a notified body.
41	BHUBANESWARI ENTERPRISES	Estar Electronic Co. Ltd.	EET-3	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number.

Sl.	Name of the vendor	Quoted Make	Quoted Model	Status of Bidder with reason of rejection
42	INLINE MEDICARE	Hebel Evidence Based Medical Technology Co. Ltd.	X6	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of valid CE certificate issued from a notified body.
43	NOTE BOOK GALLERY	CP Plus	GP 200	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of (i) Valid ISO 13485 certificate. (ii) Valid CE certificate issued from a notified body.
44	FORTUNE ENTERPRISES	Estar Electronic Co. Ltd.	EET-3B	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of Valid ISO 13485 certificate.
45	RAMSON ESTATE PVT. LTD.	Longnan Renzhong Medical Equipment Co. Ltd.	YK001	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of (i) Valid ISO 13485 certificate. (ii) Valid CE certificate issued from a notified body.
46	SATISH PHARMA	Guangzhou Berrcom Medical Device Co. Ltd.	JXB 178	<b>NOT QUALIFIED</b> because the quoted model is NOT USFDA approved with 510K number.
47	MOTIVE CONSULTANCY SERVICES	CP PLUS	CP MED	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of (i) Valid ISO 13485 certificate. (ii) Valid CE certificate issued from a notified body.
48	KANTHALI TRADERS	Wenzhou Yosun Medical Technology Co. Ltd.	YI-400	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number.
49	JYOTI PHARMACEUTICALS	Taidoc Technology Corporation	TD-1242	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number.
50	BIOMED INDIA	Hunan Honggao Electronic Technology Co. Ltd.	HG 01	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number.
51	JAIN PLASTICS	Jain Plastic	EI20JP	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of (i) Valid ISO 13485 certificate. (ii) Valid CE certificate issued from a notified body.

Sl.	Name of the vendor	Quoted Make	Quoted Model	Status of Bidder with reason of rejection
52	KHATTAR EXPORTS	Hebel Evidence Based Medical Technology Co. Ltd.	X6	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of valid CE certificate issued from a notified body.
53	EXIM LOGISTICS PVT. LTD.	KAIYI	KY 111	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of (i) Valid ISO 13485 certificate. (ii) Valid CE certificate issued from a notified body.
54	SAI ROADLINES	Henan Bingzun Industrial Co. Ltd.	BZ-R6	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of (i) Valid ISO 13485 certificate. (ii) Valid CE certificate issued from a notified body.
55	R.K.TRADERS	Raj Thermometers	NA	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of (i) Valid ISO 13485 certificate. (ii) Valid CE certificate issued from a notified body.
56	ANANTA PHARMACEUTICALS AND CHEMICALS	Manf.- Jiangxi Dayou Medical Technology, China	NX-2000	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of (i) Valid ISO 13485 certificate. (ii) Valid CE certificate issued from a notified body.
57	UTKAL DIGITAL	Manf.- Harbin Xiande Technology Development, China CP Plus	GP-200	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of (i) Valid ISO 13485 certificate. (ii) Valid CE certificate issued from a notified body.
58	Cova Medicare Pvt. Ltd.	Shenzhen Hezhizhou Technology Co.Ltd.	HT 668	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of (i) Valid ISO 13485 certificate. (ii) Valid CE certificate issued from a notified body.



Sl.	Name of the vendor	Quoted Make	Quoted Model	Status of Bidder with reason of rejection
59	DEEPAK ENTERPRISES	Huanghua Faithful Instrument Co. Ltd.	Faithful- JZK 601	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of (i) Valid ISO 13485 certificate. (ii) Valid CE certificate issued from a notified body.
60	ACE MEDICAL	Shenzhen Zheng Kang Technology Co.Ltd.	JZK 601	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of (i) Valid ISO 13485 certificate. (ii) Valid CE certificate issued from a notified body.
61	SHREE SEWA SADAN	Huizhou Xiaoou Technology Co. Ltd.	IR 988	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of valid CE certificate issued from a notified body.
62	Onkar Dies Pvt. Ltd.	Deluxe Scientific Surgico Pvt. Ltd.	DITH 120	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of valid CE certificate issued from a notified body.
63	PINAKIN IMPEX	Shenzhen Leevis Technology Co. Ltd.	TG 8818B	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of valid CE certificate issued from a notified body.
64	ARCTIC TELECOMMUNICATION PVT. LTD.	Hunan Honggao Electronic Technology Pvt. Ltd.	HG 01	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of valid CE certificate issued from a notified body.
65	SOLUMED (I)	Hunan Honggao Electronic Technology Pvt. Ltd.	HG 01	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of valid CE certificate issued from a notified body.
66	IMAGINE TECHNOLOGYS	Hunan Honggao Electronic Technology Pvt. Ltd.	HG 03	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of valid CE certificate issued from a notified body.

Sl.	Name of the vendor	Quoted Make	Quoted Model	Status of Bidder with reason of rejection
67	MAF CLOTHING PVT. LTD.	Innovative Infratech Solutions Pvt. Ltd.	C 05	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of (i) Valid CE certificate issued from a notified body. (ii) Required Formats, GST Registration Certificate & PAN copies as asked in the tender.
68	MAKEWAY FORMULATIONS PVT. LTD.	VCOYS Impetos Pvt. Ltd.	YHKY-2000	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of valid CE certificate issued from a notified body.
69	NAMANVIDHI WORLD WIDE PVT. LTD.	Hunan Honggao Electronic Technology Pvt. Ltd.	HG 01	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of valid CE certificate issued from a notified body.
70	ZIVAH INTERNATIONAL PVT. LTD.	Contec Medical Systems Co. Ltd.	TP 500	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of valid CE certificate issued from a notified body.
71	PREMIER SURGICAL	BPL Medical Technologies	Accu Digit F2/ 91MED419	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number.
72	CLASS ONE SYSTEMS S&T PVT. LTD.	Guangdong Stable Medical Technology Co. Ltd.	STBH-II	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number.
73	OM ASSOCIATES	CP Plus	GP 200	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number.
74	BHARAT BIOTECH	Harbin Xiande Technology Development Co. Ltd.	GP 300	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number.
75	RK STEEL SALES	Hunan Honggao Electronic Technology Pvt. Ltd.	HG 01	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number.
76	I-NOVATE INFOTECH PVT. LTD.	Not Specified	Not Specified	<b>NOT QUALIFIED</b> due to 1) Make & Model not mentioned in Format-B for which the quoted model could not be ascertained. 2) Non-submission of (i) Valid ISO 13485 certificate. (ii) Valid CE certificate issued from a notified body. (iii) USFDA certificate with 510K number

Sl.	Name of the vendor	Quoted Make	Quoted Model	Status of Bidder with reason of rejection
77	NATIONAL ENTERPRISERS	Shenzhen Zheng Kang Technology Co.Ltd.	JZK 601	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number.
78	ABHILASHA COMMERCIAL PVT. LTD.	Guizhou Tekang Electronics Co. Ltd.	T 168	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of (i) Valid ISO 13485 certificate. (ii) Valid CE certificate issued from a notified body.
79	PARC SOLUTIONS	Hunan JRT Medical Devices Co. Ltd.	JRT 018	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of (i) Valid ISO 13485 certificate. (ii) Valid CE certificate issued from a notified body.
80	TURQUOISE TRADELINE AND LOGISTICS	Hunan Tuogao Medical Technology Co. Ltd.	TG 8818B	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of valid CE certificate issued from a notified body.
81	NARAYAN INDUSTRIES	Jiangxi Hawk Medical Supplies Co. Ltd.	HK-T1	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of (i) Valid ISO 13485 certificate. (ii) Valid CE certificate issued from a notified body.
82	SAP INFRA TECHNOLOGIES PVT. LTD.	CP Plus	YS-ET 05	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of (i) Valid ISO 13485 certificate. (ii) Valid CE certificate issued from a notified body.
83	JR BIOTECH	Saiswaram Trade links	NA	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of (i) Valid ISO 13485 certificate. (ii) Valid CE certificate issued from a notified body.
84	MARVEL VINYLs LTD.	Jziki	JZK-601	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of valid CE certificate issued from a notified body.

Sl.	Name of the vendor	Quoted Make	Quoted Model	Status of Bidder with reason of rejection
85	PRINCE ENTERPRISES AND CO.	QQZM	GP 300	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of (i) Valid ISO 13485 certificate. (ii) Valid CE certificate issued from a notified body.
86	SHREE GOPAL SURGICALS AND MEDICINES	CP MED	BSX 815	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of (i) Valid ISO 13485 certificate. (ii) Valid CE certificate issued from a notified body.
87	PENTAGON AUTOMATION CONCEPTS	Herbin Xiande Technology Development Co. Ltd.	GP 200	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number.
88	TASHI INFOCOM PVT. LTD.	Shenzhen Quick Zoom Tech Co. Ltd.	GP 300	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of (i) Valid ISO 13485 certificate. (ii) Valid CE certificate issued from a notified body.
89	SHIVOHAM ENTERPRISES	Guangzhou Berrcom Medical Device Co. Ltd.	JXB 178	<b>NOT QUALIFIED</b> because the quoted model is NOT USFDA approved with 510K number.
90	SAIDEEP SECURITY SYSTEM AND SOLUTION	Shenzhen Yostand Technology Co. Ltd.	YS-ET03	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of (i) Valid ISO 13485 certificate. (ii) Valid CE certificate issued from a notified body.
91	WAREHOUSE INTERNATIONAL PVT. LTD.	Guangzhou Berrcom Medical Device Co. Ltd.	JXB 178	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of (i) Valid ISO 13485 certificate. (ii) Valid CE certificate issued from a notified body.
92	CHANDRAVARSHA TRADER PVT. LTD.	Herbin Xiande Technology Development Co. Ltd.	GP 300	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number.
93	HEALTH PHARMACEUTICAL	Leelvis	TG 8818B	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number.

Sl.	Name of the vendor	Quoted Make	Quoted Model	Status of Bidder with reason of rejection
94	VITA UDYOG	Microtek	TG 8818C	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of (i) Valid ISO 13485 certificate. (ii) Valid CE certificate issued from a notified body.
95	TOPSIA ESTATES PVT. LTD.	Huizhou Xiaouu Technology Co. Ltd.	IR 988	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of (i) Valid ISO 13485 certificate. (ii) Valid CE certificate issued from a notified body.
96	H.Q. LAMPS MFG CO. PVT. LTD.	HQ Lamps Manufacturing Co. Pvt. Ltd.	HQ 8801	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of valid CE certificate issued from a notified body.
97	ABJA LED PVT. LTD.	DINKANG	HG 03	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of (i) Valid ISO 13485 certificate. (ii) Valid CE certificate issued from a notified body.
98	MOHAPATRA SCIENTIFIC SUPPLY SYNDICATE	Avantor	IBSANX2000	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of (i) Valid ISO 13485 certificate. (ii) Valid CE certificate issued from a notified body.
99	NEELIMA OVERSEAS MULTIPURPOSE PVT LTD	HTC	HTC SCAN II	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of (i) Valid ISO 13485 certificate. (ii) Valid CE certificate issued from a notified body.
100	ROSMERTA TECHNOLOGIES LTD.	Hunan Honggao Electronic Technology Pvt. Ltd.	HG 03	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number.
101	TRISIM GLOBAL SOLUTION LLP	CP PLUS (Zhengzhou AiQURA Intelligent Technology Co. Ltd. )	GP 200	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of (i) Valid ISO 13485 certificate. (ii) Valid CE certificate issued from a notified body.

Sl.	Name of the vendor	Quoted Make	Quoted Model	Status of Bidder with reason of rejection
102	PLANET PHARMA	Estar Electronic Co. Ltd.	EET-3	<b>NOT QUALIFIED</b> because the quoted model is NOT USFDA approved with 510K number.
103	PANACHA INNOVATION LTD.	YIIONE	YRK-002A	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of (i) Valid ISO 13485 certificate. (ii) Valid CE certificate issued from a notified body.
104	SANDOR MEDICAIDS PVT. LTD.	Vphore Labs Pvt. Ltd.	vThermo-IE7A	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of valid CE certificate issued from a notified body.
105	IMPRESSION	Zhangye Nuonan Medical Equipment Co. Ltd.	T-01	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number.
106	S S MASER TECHNOLOGY PVT. LTD.	Shenzhen Zhengkang Technology Co.Ltd.	JZK-601	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of Valid ISO 13485 certificate.
107	INNOVATIVIEW	CP PLUS	SK-T008	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of valid CE certificate issued from a notified body.
108	CONVERGENT ALLIANCE	CP PLUS	SK-T008	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of valid CE certificate issued from a notified body.
109	ALLIANCE TRADE LINKS	Hunan Tuogao Medical Technology Co. Ltd.	TG 8818N	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of (i) Valid ISO 13485 certificate. (ii) Valid CE certificate issued from a notified body.
110	JN WORLDWIDE	XNO	(Multiple model quoted) HS 8898/ HT 668	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of valid CE certificate issued from a notified body. 3) Multiple models quoted.

Sl.	Name of the vendor	Quoted Make	Quoted Model	Status of Bidder with reason of rejection
111	KRISHNA TELENET PVT. LTD.	Dongguan HCS Electronics Co. Ltd.	HET-R121	<b>NOT QUALIFIED</b> because 1) Non-submission of Valid CE certificate issued from a notified body. 2) The furnished ISO 13485 certificate has expired.
112	JAI MATADI PHARMA	Shenzhen Aoj Medical Technology	AOJ-20A	<b>NOT QUALIFIED</b> because the quoted model has deviation in the technical specification, i.e. as per the USFDA certificate, the measurement distance is 0 cm instead of 5 cm or better..
113	LABEX KK INTERNATIONAL	Dongguan SIZMO Electronic Technology Co. Ltd.	HW-F7	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of Valid CE certificate issued from a notified body.
114	SRI SAI BALAJI ENTERPRISES	Joytech Health Care Co. Ltd (SEJOY)	DET-306	<b>NOT QUALIFIED</b> because the quoted model has deviation in the technical specification, i.e. as per the USFDA certificate, the measurement distance is 2 to 3 cm instead of 5 cm or better.
115	ZOROY ELECTRONICS LTD.	ZOROY / ZIL / IRT/3003	IASIN50	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of (i) Valid ISO 13485 certificate. (ii) Valid CE certificate issued from a notified body.
116	EMBLEM LUGGAGE COMPANY	KZED Guizhou Xinhui Optoelectronic Technology Co.	KZED-8801	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of (i) Valid ISO 13485 certificate. (ii) Valid CE certificate issued from a notified body.
117	ANAND INTERNATIONAL	Jiangxi Kinkyu Medical Technology Co. Ltd.	EW-801	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of Valid CE certificate issued from a notified body.
118	PRAGMATIC INDUSTRIAL AUTOMATION SOLUTIONS	Infrared Cameras Inc (Texas, USA)	HotSpot IR	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of (i) Valid ISO 13485 certificate. (ii) Valid CE certificate issued from a notified body.
119	MEDILINK INFO SERVICES	Joytech Health Care Co. Ltd	DET-306	<b>NOT QUALIFIED</b> because the quoted model has deviation in the technical specification, i.e. as per the USFDA certificate, the measurement distance is 2 to 3 cm instead of 5 cm or better.

Sl.	Name of the vendor	Quoted Make	Quoted Model	Status of Bidder with reason of rejection
120	EPITOM ELECTRONICS(I) PVT. LTD.	Shenzhen Aoj Medical Technology	F-103	<b>NOT QUALIFIED</b> because the quoted model is NOT USFDA approved with 510K number.
121	CELKON IMPEX PVT. LTD.	CELKON IMPEX PVT. LTD.	IR/CH-004	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of Valid CE certificate issued from a notified body.
122	SILVERLINE TECHNOLOGIES	ZHENGZHOU AiQURA intelligent Technology Connect Pvt Ltd	AD801	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of (i) Valid ISO 13485 certificate. (ii) Valid CE certificate issued from a notified body.
123	SABIR IMPEX	Shenzhen Finicare Co Ltd	FC-IR100	<b>NOT QUALIFIED</b> because the quoted model has deviation in the technical specification, i.e. as per the USFDA certificate, the measurement distance is 0 cm instead of 5 cm or better.
124	HARIKRUPA CHEMIST	MAIZA	F1	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of (i) Valid ISO 13485 certificate. (ii) Valid CE certificate issued from a notified body. (iv) GST Registration Certificate & PAN copies as asked in the tender. (v) Authorization certificate from OEM or importer.
125	AMANAT INFRASTRUCTURE PVT. LTD.	Format B Not submitted but from furnished Authorization Make: Hunan Tuogao Medical Technology Co. Ltd, may be assumed.	Format B Not submitted but from furnished & brochure Model: TG8818C may be assumed.	<b>NOT QUALIFIED</b> due to Non-submission of (i) Format-A & B as asked in the tender for which the quoted model could not be ascertained. (ii) Valid ISO 13485 certificate. (iii) Valid CE certificate issued from a notified body. (iv) USFDA certificate with 510K number
126	PROTONIC MEDICAL SYSTEMS PVT. LTD.	Format B Not submitted but from furnished Authorization Make: Protonic may be assumed.	Not Known	<b>NOT QUALIFIED</b> due to Non-submission of Valid CE certificate issued from a notified body, USFDA Certificate with 510k number, Format A & B as asked in the tender.
127	RENU RECYCLING CO.	Henan BingZun Industrial Co. Ltd	BZ-R3	<b>NOT QUALIFIED</b> due to Non-submission of Valid CE certificate issued from a notified body, GST Registration Certificate & PAN copies as asked in the tender.



Sl.	Name of the vendor	Quoted Make	Quoted Model	Status of Bidder with reason of rejection
128	HYDROGEN INFOCOM	Contec	TP500	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of (i) Valid ISO 13485 certificate. (ii) Valid CE certificate issued from a notified body.
129	BMR DISTRIBUTOR PVT. LTD.	ELECXEC Global Ltd	Non Contact Infrared Thermometer	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of (i) Valid ISO 13485 certificate. (ii) Valid CE certificate issued from a notified body.
130	SWADESH SYSTEM PVT. LTD.	Realtime	JRT 018	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of (i) Valid ISO 13485 certificate. (ii) Valid CE certificate issued from a notified body. (iii) Authorization certificate from OEM or importer.
131	ITIZEN	Zonerich	T2020	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of (i) Valid ISO 13485 certificate. (ii) Valid CE certificate issued from a notified body.
132	FINDOC INVESTMART PVT. LTD. (FD FINDOC)	Yibin Junxin Electronic Technology Co. Ltd.	T8	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of (i) Valid ISO 13485 certificate. (ii) Valid CE certificate issued from a notified body.
133	IMS MERCANTILES PVT. LTD.	Digitek / IMS Mercantiles Pvt Ltd.	HG 03	<b>NOT QUALIFIED</b> because the quoted model is NOT USFDA approved with 510K number.
134	MDAAC INTERNATIONAL PVT.LTD.	Guangzhou Berrcom Medical Device Co. Ltd	JXB 178	<b>NOT QUALIFIED</b> because the quoted model is NOT USFDA approved with 510K number.
135	AKAAR BUSINESS	Make is not mentioned in Format B	Model is not mentioned in Format B	<b>NOT QUALIFIED</b> because 1) USFDA certificate with 510K number not submitted. 2) Non-submission of Valid CE certificate issued from a notified body.
136	GROME MEDICAL	Zhejiang Mondial Electronic Technology Co Ltd.	MD 33520	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of Valid CE certificate issued from a notified body.

Sl.	Name of the vendor	Quoted Make	Quoted Model	Status of Bidder with reason of rejection
137	FOSTER MEDICAL DEVICES PVT. LTD.	ZHENGZHOU AiQURA intelligent Technology Connect Pvt Ltd (Zebronics)	AD801	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of (i) Valid ISO 13485 certificate. (ii) Valid CE certificate issued from a notified body.
138	PUNTA INDIA PVT. LTD.	Henan BingZun Industrial Co. Ltd (Sunphor)	BZ-R6	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of (i) Valid ISO 13485 certificate. (ii) Valid CE certificate issued from a notified body.
139	G B VENTURES MEDICARE DIVISION	Shenzhen Everbest Machinery Industries Co Ltd (BYD)	DT 8806S	<b>NOT QUALIFIED</b> because the quoted model is NOT USFDA approved with 510K number.
140	MAGTECH SECURITY SYSTEMS PVT. LTD.	Shenzhen ZhengKang Technology Co. Ltd (JZIKI)	JZK-601	<b>NOT QUALIFIED</b> because the quoted model is NOT USFDA approved with 510K number.
141	RHINO CONTINGS INDIA	ESTAR ELECTRONICS Co. Ltd	EET-3	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of Valid CE certificate issued from a notified body.
142	LOFTY OPTICAL INDUSTRIES	Zhangye Nuoan Medical Equipment Co. Ltd	(Multiple model quoted) Thermometer T-01 T-02 T-03	<b>NOT QUALIFIED</b> because 1) Multiple models quoted 2) The quoted model is NOT USFDA approved with 510K number.
143	BOMBAY TOOLS	Dongguan Dunniao Electronic Technology Co. Ltd	DN-998	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of (i) Valid ISO 13485 certificate. (ii) Valid CE certificate issued from a notified body.
144	AMTRAK TECHNOLOGIES PVT. LTD.	Secureeye (MediSec)	T-168	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of Valid CE certificate issued from a notified body.
145	ALLURE ENTERPRISES PVT. LTD.	ESTAR ELECTRONICS Co. Ltd	EET-3	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of Valid CE certificate issued from a notified body.
146	P. BHOGILAL PVT. LTD.	Shenzhen Aeon Technology Co. Ltd	A200	<b>NOT QUALIFIED</b> because the quoted model has deviation in the technical specification, i.e. as per the USFDA certificate, the measurement distance is less than 3 cm instead of 5 cm or better.

Sl.	Name of the vendor	Quoted Make	Quoted Model	Status of Bidder with reason of rejection
147	ZETFLY ONLINE SERVICES	Jiangsu Yuyue Medical Equipment & Supply Co. Ltd	YT-1	<b>NOT QUALIFIED</b> because the quoted model is NOT USFDA approved with 510K number.
148	K G N TECHNOLOGIES	Beila (GZ) Intelligent Information Technology Co. Ltd. / Guizhou Xinhui Optoelectronic Tech Co. Ltd	(Multiple model quoted) CK-T1501 / KZED-8801	<b>NOT QUALIFIED</b> because 1) Multiple model quoted. 2) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of (i) Valid ISO 13485 certificate. (ii) Valid CE certificate issued from a notified body. (iii) Authorization certificate from OEM or importer.
149	SHREETEX COMMERCIAL PVT. LTD.	Shenzhen Weird Power Technology Ltd. as per format B	(Multiple model quoted) KY-111 / HT 668	<b>NOT QUALIFIED</b> because of 1) Non-submission of Valid CE certificate issued from a notified body. 2) Multiple model quoted
150	MEVISION TECHNOLOGIES PVT. LTD.	Format B Not submitted	Format B Not submitted	<b>NOT QUALIFIED</b> due to Non-submission of (i) Format- B as asked in the tender for which the quoted model could not be ascertained. (ii) Valid ISO 13485 certificate. (iii) Valid CE certificate issued from a notified body. (iv) USFDA certificate with 510K number. (iv) GST Registration Certificate & PAN copies, signed tender notice with Annexure-A and Format-A, B, C as asked in the tender.
151	FOREVER YOURS RETAIL PVT. LTD.	Wisepower Enterprise (i) Pvt Ltd	Not Mentioned in Format B.	<b>NOT QUALIFIED</b> due to 1) Model not mentioned in Format-B for which the quoted model could not be ascertained. 2) Non-submission of (i) Valid ISO 13485 certificate. (ii) Valid CE certificate issued from a notified body. (iii) USFDA certificate with 510K number
152	NIGAM SUPPLIERS	Harbin Xiande Technology Development Co. Ltd (CP-Med)	GP-200	<b>NOT QUALIFIED</b> because of Non-submission of Valid ISO 13485 certificate.
153	S.V HEALTHCARE SOLUTIONS	JBDS POWER	T 7861	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of (i) valid CE certificate issued from a notified body. (ii) Authorization certificate from OEM or importer.
154	IVTEX CORP SOLUTIONS PVT. LTD.	Manf.- Harbin Xiande Technology Development, China	GP 200	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number.

Sl.	Name of the vendor	Quoted Make	Quoted Model	Status of Bidder with reason of rejection
155	LARANYA ENTERPRISES	Huizhou Xiaouu Technology Co. Ltd.	IR 988	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of valid CE certificate issued from a notified body.
156	SACAYA TECHNOLOGIES PVT. LTD.	Henan Bingzun Industrial Co. Ltd.	YS-ET05	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of (i) Valid ISO 13485 certificate. (ii) Valid CE certificate issued from a notified body.
157	IPRAVARA TECHNOLOGY AND CONSULTANCIES	Green Ocean Research Labs Pvt. Ltd.	ASHA Fit	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of (i) Valid ISO 13485 certificate. (ii) Valid CE certificate issued from a notified body.
158	GREENTEC CONSTRUCTION	Hangzhou Universal Electronic Co. Ltd.	EF-100A	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number.
159	RASHMI ASSOCIATES	Guizhou Tekang Electronics Co. Ltd.	T-168	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of (i) Valid ISO 13485 certificate. (ii) Valid CE certificate issued from a notified body.
160	INNOVATIVE IDEALS & SERVICES (I) LTD.	Global Communication	GLB 2000	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of (i) Valid ISO 13485 certificate. (ii) Valid CE certificate issued from a notified body.
161	SWAPU PACKAGING	Manf.- Harbin Xiande Technology Development, China CP Plus	GP 200	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number.
162	A.R. INTERNATIONAL	Guangzhou Nanxue Medical Instruments Co. Ltd.	NX-2000, HY-01	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of (i) Valid ISO 13485 certificate. (ii) Valid CE certificate issued from a notified body.
163	SATYAM HEALTHCARE	Hunan Tuogao Medical Technology Co. Ltd.	TG 8818B	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of Valid CE certificate issued from a notified body.

Sl.	Name of the vendor	Quoted Make	Quoted Model	Status of Bidder with reason of rejection
164	BISWAL TRADE LINK PVT. LTD.	Bo-HUI	T-168	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of (i) Valid ISO 13485 certificate. (ii) Valid CE certificate issued from a notified body.
165	LTB INTRACONSULTANT PVT. LTD.	Henan Yobekan Medical Equipment Co. Ltd.	Wplus	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of (i) Valid ISO 13485 certificate. (ii) Valid CE certificate issued from a notified body.
166	ZINOTA REMEDIES PVT. LTD.	Guangzhou Berrcom Medical Devices Co. Ltd.	JXB 178	<b>NOT QUALIFIED</b> because the quoted model is NOT USFDA approved with 510K number.
167	ORISSA TRADING CORPORATION LTD.	Henan Bingzun Industrial Co. Ltd.	BZ-R6	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of valid CE certificate issued from a notified body.
168	KWALITY PHOTONICS PVT. LTD.	Medisec	T-168	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of (i) Valid ISO 13485 certificate. (ii) Valid CE certificate issued from a notified body.
169	MEDI PHARMA	Dongguan Simzo Electronic Technology Co. Ltd.	HW-F7	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number.
170	BHAWANI MEDICAL AGENCIES	Dongguan Simzo Electronic Technology Co. Ltd.	HW-F7	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number.
171	ADWELL INTERNATIONAL PVT. LTD.	Zhengzhou Aiqura Intelligent technology Co. Ltd.	AD801	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of (i) Valid ISO 13485 certificate. (ii) Valid CE certificate issued from a notified body.
172	ADPRINT INTERNATIONAL	Harbin Xiande Technology Development, China	GP 300	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number.

Sl.	Name of the vendor	Quoted Make	Quoted Model	Status of Bidder with reason of rejection
173	SARALA ENTERPRISES	Hunan Zhongliankang Medical Technology Co. Ltd.	ZLK-IRT101	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of (i) Valid ISO 13485 certificate. (ii) Valid CE certificate issued from a notified body.
174	NEHA LIFESCIENCE PVT. LTD.	Harbin Xiande Technology Development, China	GP 200	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number.
175	ELEGANT CARR JEWELS	Alicn Medical	(Multiple model quoted) AET-R731/ R1B1/ R1F1/ R1D2	<b>NOT QUALIFIED</b> due to 1) Non-submission of Valid ISO 13485 certificate. 2) Multiple model quoted.
176	SHREE SHYAM TECHINDIA PVT. LTD.	Alicn Medical	(Multiple model quoted) AET-R731/ R1B1/ R1F1/ R1D1	<b>NOT QUALIFIED</b> due to 1) Non-submission of Valid ISO 13485 certificate. 2) Multiple model quoted.
177	PTC ENGINEERING (I) PVT. LTD.	Jiangxi Dayou Medical Technology Co. Ltd.	IR Thermometer	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of (i) Valid ISO 13485 certificate. (ii) Valid CE certificate issued from a notified body.
178	MAFATLAL INDUSTRIES LTD.	Changshan Estar Electronics Co. Ltd.	EET-3	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number.
179	HLBS TECH PVT. LTD.	Hunan Aodun Electronic Technology Co. Ltd.	AD 801	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number.
180	NEW GIFT INN	Dongguan Simzo Electronic Technology Co. Ltd.	HW-F7	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number.
181	GREENCO SYSTEM PVT. LTD.	Manf.- Harbin Xiande Technology Development, China CP Plus	GP 200	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number.
182	INTERNATIONAL TRADE LINK	Hunan Honggao Electronic Technology Co. Ltd.	HG 01	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of valid CE certificate issued from a notified body.
183	J.K.ASSOCIATES	Manf.- Harbin Xiande Technology Development, China CP Plus	GP 200	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number.

Sl.	Name of the vendor	Quoted Make	Quoted Model	Status of Bidder with reason of rejection
184	MEDI PLUS	Guangzhou Berrcom Medical Devices	JXB-182 (K130231)	<b>NOT QUALIFIED</b> because the quoted model has deviation in the technical specification, i.e. as per the USFDA certificate, Accuracy of measurement is $\pm 0.3^{\circ}$ C instead of $0.2^{\circ}$ C/ $^{\circ}$ F or better.
185	SOUTH INDIA SURGICAL (I) PVT. LTD.	Penrui Biotech Inc.	JRT 200	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of valid CE certificate issued from a notified body.
186	SCHEME ART REAL ESTATE PVT. LTD.	Microtek International Pvt. Ltd.	IT-1520	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of valid CE certificate issued from a notified body.
187	TKG SCIENTIFIC INSTRUMENTS	Surezen Hebei Evidence Based Medical Technology Co. Ltd.	X6	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of (i) Valid ISO 13485 certificate. (ii) Valid CE certificate issued from a notified body.
188	INDIUM TECH SPIRE PVT. LTD.	Vantage	VV-F-168-L5V	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of (i) Valid ISO 13485 certificate. (ii) Valid CE certificate issued from a notified body.
189	MANISHA LIFESPACES	Not Mentioned in Format B.	Not Mentioned in Format B.	<b>NOT QUALIFIED</b> because 1) Make and model not mentioned in the prescribed format B. Hence, the quoted model could not be ascertained.
190	MANISHA LIFESPACES CONSULTING PVT LTD	Not Mentioned in Format B.	Not Mentioned in Format B.	<b>NOT QUALIFIED</b> because 1) Make and model not mentioned in the prescribed format B. Hence, the quoted model could not be ascertained.
191	VINAY NITRATES	Not Mentioned in Format B.	Not Mentioned in Format B.	<b>NOT QUALIFIED</b> because 1) Make and model not mentioned in the prescribed format B. 2) From the furnished USFDA & CE certificates, The models mentioned in USFDA certificate having 510K number and that mentioned in the CE certificate is not the same. Hence the quoted model could not be ascertained.

Sl.	Name of the vendor	Quoted Make	Quoted Model	Status of Bidder with reason of rejection
192	BLG INTERNATIONAL PVT. LTD.	Bing Zun	BZ-R6	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of valid CE certificate issued from a notified body.
193	ZERO 1 TEKTRONICS LLP	Zeroi Tektronics LLP	TG 01	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of (i) Valid ISO 13485 certificate. (ii) Valid CE certificate issued from a notified body.
194	COSMAS RESEARCH LAB LTD.	Huizhou Xiaoou Technology Co. Ltd.	IR 988	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of valid CE certificate issued from a notified body.
195	SG MEDICARE PVT. LTD.	Estar Electronics Co. Ltd.	EET-3	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number.
196	LARS MEDICARE PVT. LTD.	Lars Medicare	Not Specified	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of valid CE certificate issued from a notified body.
197	DIGITAL SHARK TECHNOLOGY PVT. LTD.	Hunan Honggao Electronic Technology Co. Ltd.	HG 01	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of valid CE certificate issued from a notified body.
198	TIRUMALA INFRASTRUCTURE & DEVELOPMENT PVT. LTD.	Vizin India Pvt. Ltd.	VOX-1233	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of valid CE certificate issued from a notified body.
199	UNITED ASSOCIATES	KZED-8801	KZED-8801	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of (i) Valid ISO 13485 certificate. (ii) Valid CE certificate issued from a notified body.
200	MEDILLOYD MEDICAMEN PVT. LTD.	DHRUVH IMPEX	YT-EWQ-04	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of (i) Valid ISO 13485 certificate. (ii) Valid CE certificate issued from a notified body.



Sl.	Name of the vendor	Quoted Make	Quoted Model	Status of Bidder with reason of rejection
201	GLADIOLUX ENERGY SOLUTIONS PVT. LTD.	Hangzhou Qingyuan Medical Equipment Technology co.ltd	QY-EWQ-01	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of (i) Valid ISO 13485 certificate. (ii) Valid CE certificate issued from a notified body.
202	UNITED ENTERPRISES	Microtek	TG8818C	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of (i) Valid ISO 13485 certificate. (ii) Valid CE certificate issued from a notified body.
203	NAMAN ENTERPRISES	Pingxiang Orange Technology Co.Ltd	ZG-H01	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of (i) Valid ISO 13485 certificate. (ii) Valid CE certificate issued from a notified body.
204	ASPIRE PHARMACEUTICALS	Harbin Xiande Technology Development Co.Ltd	GP-300	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of valid CE certificate issued from a notified body.
205	INNERVISION SYSTEMS	Medisec By Secureye	T-168	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of (i) Valid ISO 13485 certificate. (ii) Valid CE certificate issued from a notified body.
206	DEEPAK KUMAR AGARWALS	Hangzhou Qingyuan Medical Equipment Technology	Qy-EWQ-01	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of (i) Valid ISO 13485 certificate. (ii) Valid CE certificate issued from a notified body.
207	PRAGATI CONSTRUCTIONS	CP-Med(Harbin Xlande Technology Development CoLtd)	GP-200	<b>NOT QUALIFIED</b> due to Non-submission of Valid CE certificate issued from a notified body.
208	UTKAL COMPUTER CARE	CP-Med(Harbin Xlande Technology Development CoLtd)	GP-200	<b>NOT QUALIFIED</b> due to Non-submission of Valid CE certificate issued from a notified body.
209	VIKRANT LIFE SCIENCES PVT. LTD.	Guangzhou AXD Electronic Co.Ltd	AXD-515	<b>NOT QUALIFIED</b> due to Non-submission of Valid CE certificate issued from a notified body.

Sl.	Name of the vendor	Quoted Make	Quoted Model	Status of Bidder with reason of rejection
210	DFS BUISNESS SOLUTION	Alphamed	UFR106	<b>NOT QUALIFIED</b> because the quoted model is NOT USFDA approved with 510K number.
211	AXESS MEDICOM PVT LTD	1)Hunan Honggao electronic technology co.ltd 2) Shenzhen AOJ Medical Technology Co.Ltd.	(Multiple model quoted) HG01 / F103	<b>NOT QUALIFIED</b> because (i) Multiple model quoted (ii) The quoted models are NOT USFDA approved with 510K number.
212	AGMATEL (I) PVT. LTD.	Guangzhou Berrcom medical device co.ltd	JXB-178	<b>NOT QUALIFIED</b> because the quoted model is NOT USFDA approved with 510K number.
213	BIDYASHREE ELECTRONICS PVT. LTD.	Hongyi JIG Rapid Technologies	Not mentioned in the format-B	<b>NOT QUALIFIED</b> because 1) The model is NOT USFDA approved with 510K number. 2) Non-submission of (i) Valid ISO 13485 certificate. (ii) Valid CE certificate issued from a notified body.
214	LN WRENCH	Dongguan SIMZO Electronic Technology	HW-F7	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of (i) Valid ISO 13485 certificate. (ii) Valid CE certificate issued from a notified body. (iii) Authorization certificate from OEM or Importer.
215	VARENYAM HEALTHCARE PVT. LTD.	Hunan Yinghe Kangyvan biological engineering ltd	YHKY-2000	<b>NOT QUALIFIED</b> due to Non-submission of (i) Valid ISO 13485 certificate. (ii) Valid CE certificate issued from a notified body.
216	KANHA ENTERPRISES	Harbin Xiande Technology Development Co.Ltd	GP-300	<b>NOT QUALIFIED</b> due to Non-submission of Valid CE certificate issued from a notified body.
217	ADITYA ENTERPRISES	Jain Plastics	Taapman - EI20JP	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of valid CE certificate issued from a notified body.
218	SRISRI TRADE CONTRACT	Vizin India Pvt. Ltd.	ARU-MED	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of valid CE certificate issued from a notified body.
219	IZAARA OVERSEAS PVT. LTD.	Alicn Medical	AET-R171	<b>NOT QUALIFIED</b> due to Non-submission of (i) Valid ISO 13485 certificate. (ii) Valid CE certificate issued from a notified body.
220	RANISH IMPEX PVT. LTD.	Xiamen Ants Bro Technology Co. Ltd.	HA-650	<b>NOT QUALIFIED</b> due to Non-submission of valid ISO 13485.

Sl.	Name of the vendor	Quoted Make	Quoted Model	Status of Bidder with reason of rejection
221	SYNERGY DIAGNOSTICS PVT. LTD.	Hunan Honggao Electronic Technology Co. Ltd.	HG 01	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of (i) Valid ISO 13485 certificate. (ii) Valid CE certificate issued from a notified body.
222	NOVATIUM SOLUTIONS PVT LTD	Henan Bingzun Industrial Co. Ltd.	BZ-R6	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of (i) Valid ISO 13485 certificate. (ii) Valid CE certificate issued from a notified body.
223	SIN FYNAMICS	Hunan Honggao Electronic Technology Co. Ltd.	HG 01	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of (i) Valid ISO 13485 certificate. (ii) Valid CE certificate issued from a notified body.
224	ARESS ENTERPRISERS	Hunan Honggao Electronic Technology Co. Ltd.	DIKANG HG 01	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number.
225	SHEO AND COMPANY	Manf.- Harbin Xiande Technology Development, China CP Plus	GP-200	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number.
226	SHYAMSHREE RESIDENCY PVT. LTD.	Web kraft	EP 510	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of (i) Valid ISO 13485 certificate. (ii) Valid CE certificate issued from a notified body.
227	GOEL ENTERPRISES	Goel Enterprises	Prodot	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of (i) Valid ISO 13485 certificate. (ii) Valid CE certificate issued from a notified body.
228	TOSHNIWAL INDUSTRIES PVT. LTD.	Hetaida	HTD8813C	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number.
229	PETRYC INTERNATIONAL PVT. LTD.	Jiangxi AOV Maternity and baby Products Co. Ltd.	AOV8711	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of (i) Valid ISO 13485 certificate. (ii) Valid CE certificate issued from a notified body.
230	A.K TRADERS	Jiangxi AOV Maternity and baby Products Co. Ltd.	AOV8711	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of (i) Valid ISO 13485 certificate. (ii) Valid CE certificate issued from a notified body.

Sl.	Name of the vendor	Quoted Make	Quoted Model	Status of Bidder with reason of rejection
231	SCICHEM CORPORATION	Shenzhen Yostand Technology Co. Ltd.	YS-ET05	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of (i) Valid ISO 13485 certificate. (ii) Valid CE certificate issued from a notified body.
232	AGGRESSIVE ELECTRONICS MANUFACTURING SERVICES	Aggressive Electronics Manufacturing Services Pvt. Ltd.	Alpha T-168A	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of (i) Valid ISO 13485 certificate. (ii) Valid CE certificate issued from a notified body.
233	HUMANBIOLOGICS PRODUCT AND SERVICE PVT. LTD.	Shenzhen AOJ Medical Technology Co. Ltd.	AOJ-20A	<b>NOT QUALIFIED</b> because the quoted model has deviation in the technical specification, i.e. as per the USFDA certificate, the measurement distance is 0 cm instead of 5 cm or better.
234	EUGEN PHARMA	Yobekan	KV-11	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of (i) Valid ISO 13485 certificate. (ii) Valid CE certificate issued from a notified body.
235	SRA EQUIPMENTS PVT. LTD.	Shenzhen AOJ Medical Technology Co. Ltd.	(Multiple model quoted) AOJ-20A / AOJ-20B	<b>NOT QUALIFIED</b> because 1) The bidder has quoted multiple models. 2) The quoted model has deviation in the technical specification, i.e. as per the USFDA certificate, the measurement distance is 0 cm instead of 5 cm or better.
236	DEZAL ELECTRO INC.	BioDOT/Bo-HUI	T-168	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of (i) Valid ISO 13485 certificate. (ii) Valid CE certificate issued from a notified body. (iii) Authorization certificate from OEM or Importer. (iv) Required Formats, GST Registration Certificate & PAN copies as asked in the tender.
237	PADMAVATI INC.	LUZHOU Skinod Technology Co. Ltd.	SK T008	<b>NOT QUALIFIED</b> because 1) The quoted model is NOT USFDA approved with 510K number. 2) Non-submission of valid CE certificate issued from a notified body.
238	EMMTECH CALIBRATION			<b>NOT CONSIDERED FOR BID EVALUATION BECAUSE THE BIDS HAVE BEEN RECEIVED AFTER THE DUE DATE &amp; TIME OF BID SUBMISSION.</b>
239	FABI CORPORATION PVT LTD			
240	IDEA TECHNOSOLUTIONS PVT. LTD.			
241	IMORA CORPORATION SERVICES PVT. LTD.			
242	LKLS PVT. LTD.			
243	Morepen Lab Ltd.			
244	RAJBORASAMBAR MEDICAL HALL			
245	SPERO MEDICAL SOLUTION			

**III. LIST OF BIDDERS WHO HAD SUBMITTED THEIR SAMPLES ALONG WITH THEIR BID**

Sl.	Name of the vendor	Remarks
1	ADPRINT INTERNATIONAL	Sample received
2	ADWELL INTERNATIONAL PVT. LTD.	Sample received
3	ARCTIC TELECOMMUNICATION PVT. LTD.	Sample received
4	BISWAL TRADE LINK PVT. LTD.	Sample received
5	HYDROGEN INTOCORN	Sample received
6	INNOVATIVEVIEW	Sample received
7	NERCELLENCE CONSULTANCY SERVICES	Sample received
8	QUICK PAY PVT. LTD.	Sample received
9	RK STEEL SALES	2 Sample received
10	SHREETEX COMMERCIAL PVT. LTD.	Sample received
11	SOLUMED (I)	Sample received

**NOTE:**

- 1) The date of price bid opening is scheduled on **Dt.10-07-2020 at 03:30 PM.**
- 2) All the technically qualifying bidders are requested to be present or send authorized representatives to be present during the price bid opening.
- 3) All the bidders who have submitted samples along with their bids are requested to take back their samples from the office of the undersigned during office hours, **i.e. from 11 am to 5 pm.**

**Sd/-**  
**GM Operation (Equipment), OSMCL**