

## **CORRIGENDUM**

The Corrigendum has been made to the earlier published Pre-bid amendment point no-18 & 22 for **Ultrasound Machine with Color Doppler (Bid Ref. No.- OSMCL/2018-19/EQP-RADIOLOGY/20)** which may be read as below,

18. System should be supplied with Broad Band Linear Probe for Vascular applications having frequency range **up to 13 MHz or more.**

22. Facility of height adjustment is not mandatory.

The rest portions of the earlier published Pre-bid amendments for **Ultrasound Machine with Color Doppler** remains unchanged.

**Sd/-  
Managing Director  
OSMC**