



**Tender Reference No. : OSMCL/2018-19/EQP-ANATOMY (Retender)/12**  
**Tender Name: Retender for Supply & Installation of Equipment for Anatomy**  
**Departments of Medical Colleges**

**I. LIST OF BIDDERS & THEIR OFFERED EQUIPMENT WHICH ARE TECHNICALLY RESPONSIVE AND QUALIFIED FOR ON-LINE PRICE BID OPENING (EQUIPMENT-WISE)**

Sl.	Name of Equipment	Sl.	Name of Bidder
1	Two body Mortuary Cooler	1	Ananta Pharmaceuticals and Chemicals
		2	York Scientific Industries Pvt. Ltd.
2	Penta Head Teaching Microscope with photographic attachment	1	Lawrence and Mayo India Pvt. Ltd.
		2	Radical Scientific Equipments Pvt. Ltd.
		3	Mediplus
		4	Ananta Pharmaceuticals and Chemicals
		5	Premier Surgicals
3	Trinocular Fluroscent Microscope	1	Lawrence and Mayo India Pvt. Ltd.
		2	Mediplus
		3	Ananta Pharmaceuticals and Chemicals

**II. LIST OF BIDDERS & THEIR OFFERED EQUIPMENT WHICH ARE TECHNICALLY NON RESPONSIVE AND NOT QUALIFIED FOR PRICE BID OPENING (EQUIPMENT-WISE)**

Sl.	Name of Equipment	Name of Bidder	Reason of Rejection
1	Two body Mortuary Cooler	Mediplus, Bhubaneswar	As per para-wise compliance statement (Format T11) submitted by the bidder there is provision of two Compressor model of refrigeration system i.e. KCJ513HAE- S420H (which is ISI marked, but not CFC free) and KCJ513HAG-B420H (which is CFC free, but not ISI marked) which deviates to our tender technical specification.
		Megastar Engineering	Manufacturer authorization has been withdrawn by the manufacturer.
2	Trinocular Fluroscent Microscope	Radical Scientific Equipments Pvt. Ltd.	As per para-wise compliance statement (Format T11) submitted by the bidder the lens is Semi Apochromatic (S-Apo) instead of Apochromatic as asked in the tender technical specification.
3	Virtual Anatomy Dissection Table	Vishal Surgical Equipment Co	1. As per para-wise compliance statement (Format T11) submitted by the bidder against the availability of feature of Real time 6 axis navigation with stylus, the bidder mentioned that

			<p>“No stylus is necessary to operate the product, however if required can be provided” which is not a condition of the tender, rather the feature of stylus shall have to be with the machine.</p> <p>2. As per the tender specification the machine has to work with medical data and live person, but as per para-wise compliance (Format T11) submitted by the bidder against this, the bidder has not confirmed the working of the machine with live persons. Since working of the machine with live person is an important requirement of the specification, the suggestion of the bidder i.e. “No additional conversion necessary” cannot be considered.</p> <p>3. As per the para-wise compliance statement and technical brochure furnished the hydrophobic tempered glass is of 2 mm instead of 5 mm as asked in technical specification. Since 5mm tempered glass is a requirement of the specification for protection, the suggestion of the bidder “5mm does not add to the quality of the product, it adds to weight only” cannot be considered.</p> <p>4. As per Manufacturer authorisation (Format-T7) submitted by the bidder the authorisation document is valid until 30/04/2019 which is less than the bid validity period laid down in tender document i.e. 180 days from last date of bid submission.</p>
4	Dissecting Microscope	Lawrence and Mayo India Pvt. Ltd.	As per the brochure of OEM submitted, the working distance in the quoted model is 72 mm instead of 97 mm and illumination feature is absent instead of Ring type LED as asked in the tender technical specification.
		Radical Scientific Equipments Pvt. Ltd.	As per the technical brochure furnished by the bidder the working distance is maximum 75 mm instead of 97mm as asked in the tender technical specification. Hence the bid is technically not qualified.

**III. LIST OF BIDDERS & THEIR OFFERED EQUIPMENT WHICH ARE TECHNICALLY RESPONSIVE LEADING TO SINGLE RESPONSIVE BID**  
**(EQUIPMENT-WISE)**

<b>Sl.</b>	<b>Name of Equipment</b>	<b>Name of Bidder</b>	<b>Decision</b>
1	Embalming Machine	York Scientific Industries Pvt.Ltd.	Recommended for Price Bid Opening.
2	Virtual Anatomy Dissection Table	Maveric Solutions Inc.	Recommended for Price Bid Opening.
3	Dissecting Microscope	Premier Surgicals	Recommended for Price Bid Opening.
4	Connect Lab & Digital System	Maveric Solutions Inc.	Recommended for Price Bid Opening.

**Sd/-**  
**Managing Director**  
**OSMCL**