



**Tender Reference No. : OSMCL/2018-19/AHRCC Lab Equipments/16**  
**Tender Name: TENDER FOR SUPPLY & INSTALLATION OF LABORATORY EQUIPMENTS AT AHRCC-  
 CUTTACK & VIMSAR-BURLA**

**I. LIST OF BIDDERS & THEIR OFFERED EQUIPMENT WHICH ARE TECHNICALLY  
 RESPONSIVE AND QUALIFIED FOR ON-LINE PRICE BID OPENING  
 (EQUIPMENT-WISE)**

Sl.	Name of Equipment	Sl.	Name of Bidder
1	Semi automatic Rotary Microtome	1	MEDIAIDS
		2	VISION DIAGNOSTICS (INDIA) PRIVATE LIMITED
		3	MEDICO INTERNATIONAL
2	Automatic Tissue Processor	1	MEDIAIDS
		2	VISION DIAGNOSTICS (INDIA) PRIVATE LIMITED
		3	MEDICO INTERNATIONAL
3	Automatic Chemiluminescence Assay	1	SIEMENS HEALTHCARE PRIVATE LIMITED
		2	POCT SERVICES
4	Multi Headed Microscope with Photographic attachment	1	MEDIPLUS
		2	RADICAL SCIENTIFIC EQUIPMENTS PVT. LTD
5	Cryostat Microtome	1	MEDIAIDS
		2	VISION DIAGNOSTICS (INDIA) PRIVATE LIMITED
		3	MEDICO INTERNATIONAL
6	Cell Counter Fully Automated Analyzer (5 parts) + Retics	1	SIEMENS HEALTHCARE PRIVATE LIMITED
		2	KRISH BIOMEDICALS
7	Automatic Slide Processing and Cell Staining System	1	MEDIAIDS
		2	VISION DIAGNOSTICS (INDIA) PRIVATE LIMITED
8	Binocular Microscope (High end)	1	MEDICO INTERNATIONAL
		2	MEDIPLUS
9	Cyto Centrifuge	1	VISION DIAGNOSTICS (INDIA) PRIVATE LIMITED
		2	MEDICO INTERNATIONAL

**II. LIST OF BIDDERS & THEIR OFFERED EQUIPMENT WHICH ARE TECHNICALLY NON RESPONSIVE AND NOT QUALIFIED FOR PRICE BID OPENING (EQUIPMENT-WISE)**

Sl.	Name of Equipment	Name of Bidder	Reason of Rejection
1	Automatic Chemiluminescence Assay	MEDIAIDS	As the test throughput offered in the quoted model is 86 tests/hr instead of 100 tests/hr or more as asked in the tender technical specification.
2	Binocular Microscope (High end)	RADICAL SCIENTIFIC EQUIPMENTS PVT. LTD	In the furnished FDA certificate there is no such scope defined for the quoted model/series as available in the furnished CE certificate. Thus it is evident that the furnished FDA certificate is not compliant to the quality parameters/standards for the quoted model rather it is compliant to manufacturing quality standard.
3	Automatic Biochemical Analyzer	POCT SERVICES	Due to non-submission of EMD.

**III. LIST OF BIDDERS & THEIR OFFERED EQUIPMENT WHICH ARE SINGLE BID/RESULTED IN SINGLE BID (EQUIPMENT-WISE)**

Sl.	Name of Equipment	Name of Bidder	Decision
1	LBC (Liquid Based Cytology) Smear Preparation Instrument	VISION DIAGNOSTICS (INDIA) PRIVATE LIMITED	Recommended for retender
2	Fully Automatic IHC Slide Staining System	ROCHE DIAGNOSTICS INDIA PVT. LTD.	Recommended for retender
3	Blood Gas Analyzer	MEDIPLUS	Recommended for retender
4	Automatic Biochemical Analyzer	SIEMENS HEALTHCARE PRIVATE LIMITED	Recommended for retender
5	Fluorescent Microscope	MEDIPLUS	Recommended for retender

**III. LIST OF ITEMS FOR WHICH NO BID HAS BEEN (EQUIPMENT-WISE)**

Sl.	Name of Equipment	Decision
1	Deep Freezer (-40 Degree C)	Recommended for retender
2	Deep Freezer (-80 Degree C)	Recommended for retender
3	Gel Electrophoresis	Recommended for retender
4	Laminar Airflow Bench (Class IIa)	Recommended for retender

Sd/-  
**Managing Director**  
**OSMCL**