



**ODISHA STATE MEDICAL CORPORATION LTD.**  
**(A Govt. of Odisha Undertaking)**  
*Assuring Quality, Saving lives*

Notification No. 22.....

Date: 04/01/2022

**PRE-BID CORRIGENDUM**

**FOR SUPPLY OF SURGICAL, SUTURES, SYRINGES AND OTHER ITEMS**

**FOR THE F.Y- 2022-2023**

**Bid Ref. No.:- OSMCL/2022-23/(SUT.-SURG. AND OTHER)-DHS/04**

Sl. No	Item Name/ Particulars	Details as mentioned in the tender document	Queries raised by the prospective bidders	Clarifications/ Amendments in response to the queries.
<b>Technical Queries</b>				
1	Clause No.4.1 (Section –IV) Sl. No. 56	(S02037), Sterile Disposable Blood Lancet (Pricking Needle)	Inclusion of <b>auto disable Lancet</b>	<b>No Change</b>
2	Clause No. 5.2.16	The bidder has to furnish the declaration regarding Bid Security as per <b>Format T3 insteade of EMD</b> (Earnest Mony Deposit).	Not mentioning the EMD amount	<b>No Change</b>
3	Clause No.4.1 (Section –IV) Sl. No. 15 Polypropylene (S01050)	Size:- 2, 1/2 Circle Round Bodied 40mm (Heavy needle) U.S.P, with CE certification	<b>Size:- 2-0, 1/2 Circle Round Bodied 40mm (Heavy needle) U.S.P, with CE certification</b>	<b>No Change</b>
4	Clause No.4.1 (Section –IV) Sl. No. 20 Catgut Chromic Atraumatic (S01074)	Size:- 4-0, 1/ 2 Circle Round Bodied (45mm needle) U.S.P, with CE certification (Not Mandatory)	<b>Size:- 4-0, 3/8 Circle Round Bodied (16 mm needle) U.S.P, with CE certification (Not Mandatory)</b>	<b>No Change</b>
5	Clause No.4.1 (Section –IV) Sl. No. 23 Polypropylene (S01077)	Size:- 4-0, length: 70cm, 1/2 Circle Round Bodied 20mm (with needle) U.S.P, with CE certification	<b>Size:- 4-0, length: 70cm, 1/2 Circle Round Bodied 16mm (with needle) U.S.P, with CE certification</b>	<b>No Change</b>

Page 1



**ODISHA STATE MEDICAL CORPORATION LTD.**  
**(A Govt. of Odisha Undertaking)**  
*Assuring Quality, Saving lives*

6	Clause No.6.20.1	6.20.1 The Tender Inviting Authority's contractual right to inspect, test and, if necessary, reject the goods after the goods' arrival at the final destination shall have no bearing of the fact that the goods have previously been inspected and cleared by Tender Inviting Authority's inspector during <b>sample verification</b> .	Verify/ evaluate the sample	<b>No Change</b>
---	------------------	--	-----------------------------	------------------

**General Clarification on Price Bid**

- a) The bidders have to quote the prices as per the format mentioned in Price BOQ in the e-tender portal. The prices (Rates) are to be quoted per Each Reel/ Foil/ No./ Roll/ Piece/ Metre/ Set /Test/ Kit/ Bottle/ Jar/ Combipack/ Tube/ Pessary/ PFS/ Film etc.( as the case may be) and not in unit pack, the quoted price includes packing, forwarding / transportation & Door Delivery and excludes GST.
- b) The amendments mentioned above are to be treated as amendments in the technical specification(s) and term(s) and condition(s) of the above tender references. All other technical specifications and terms conditions remain unchanged.
- c) Since any text in the price BOQ can't be changed in the e-tender portal, the amendments mentioned in the specifications above are to be treated as amendments in the price bid/ BOQ also.

  
**GM Operations (D & S)**  
**Odisha State Medical Corporation Ltd**