



BIDDERS' SELECTION / REJECTION LIST

Notification No.: 6479

Dt.: 08.07.2021


Tender Name: e-Tender for "Supply, installation & commissioning of ECMO machine at CTVS Department of SCBMCH-Cuttack" (Bid Ref. No- OSMCL/2021-22/EQP-ECMO/04 dtd. 10.06.2021)

I. LIST OF BIDDERS & THEIR OFFERED EQUIPMENT WHICH ARE TECHNICALLY RESPONSIVE AND RECOMMENDED FOR ON-LINE PRICE BID OPENING

Sl.	Name of Equipment	Sl.	Name of Bidder
1.	EXTRA CORPOREAL MEMBRANE OXYGENATION (ECMO) MACHINE	1.	M/s MA DURGA ENTERPRISES Manufacturer- Abott, Switzerland Make- Abott Model- Centrimag
		2.	M/s B.L. LIFESCIENCES Manufacturer- Xenious AG, Germany Make- Xenious (Medos) Model- Xenious Console

II. LIST OF BIDDERS & THEIR OFFERED EQUIPMENT WHICH ARE TECHNICALLY NON RESPONSIVE & NOT RECOMMENDED FOR PRICE BID OPENING:

Sl.	Name of Bidder	Reason of rejection
1.	SAIMED INNOVATION Manufacturer- Sorin Group, Germany Make- LivaNova Model- S5 + CP5	1) Quoted system is a Heart lung machine with arrangement of ECMO function instead of a dedicated ECMO machine as per the requirement. In addition to above, the quoted system is not a compact, small and light in weight. (Not accepted by committee) 2) Has Complied for speed of the pump having Upto 3,500 rpm instead of 0 to 5,000 rpm as per the submitted compliance report (Format T10 & T11) . (Not accepted by committee) 3) Has Complied for Water reservoir capacity as per the submitted compliance report (Format T10 & T11). But has referred the technical data sheet of Oxygenator (EOS ECMO) instead of Heating Cooling Unit. In the submitted data sheet of Heater-Cooler-3T and as verified from OEM website, The model is very big with three water tanks having total 12 liters of capacity instead of 2-5 liters capacity and not a compact, small and light in weight for Ecmo system .(Not accepted by committee) 4) As per the submitted compliance report (Format T10 & T11), The quoted oxygenator have certified for 5 days use instead of The Oxygenator should be CE certified for minimum 14 days or more continuous use for Adult and Paediatric Applications.(Not accepted by committee)


08/07/2021 . Page 1 of 2

		<p>5) Has not complied Other accessories should be supplied along with the ECMO Unit as below:-</p> <p>1. Ultrasonic Gel: 6 nos.,2. Insertion kit: 6 nos.,3. Dilator: 6 nos.,4. Tubing Clamp: 12 nos., 5. Femoral Venous and Arterial Canula one each (16/17 FR ,18/19 FR,20/21 FR, 22/23 FR, 24/25 FR).</p> <p>6) Has not Complied the below consumable items for ECMO in the submitted compliance report (Format T10 & T11) as per Bid requirement:-</p> <p>I) Complete Adult ECMO Patient Kit containing Oxygenator with heat exchanger,PVC Tubing with Heparin+Albumin coated , Pump-head with Heparin+Albumin coated for Adult application.</p> <p>II) Complete Paediatric ECMO Patient Kit containing Oxygenator with heat exchanger,PVC Tubing with Heparin+Albumin coated , Pump-head with Heparin+Albumin coated for paediatric application.</p>
--	--	---

N.B: The online price bids of the qualifying bidders shall be opened on Dt. 08.07.2021 at 4.00 PM



**GM, Operation (Equipment)
Odisha State Medical Corporation Ltd.**