

NOTICE

Notice No. 771

Date 17-01-2020

This is hereby informed that OSMCL has invited Online Bids through GeM portal from eligible bidders for EIGHT (08) items related to Laboratory department.

The detailed technical specifications, terms and conditions have been defined in the bid document of GeM but for few items, product and manufacturer quality standards have not been defined. Thus, to ensure quality, the same have been specified herewith and the bidders must abide with the specified product and manufacturer quality as mentioned below;

The details of the requirements mentioned below must be complied by the prospective bidders and should quote accordingly.

1. **Name of the item & Qty** : Automated Hematology Analyser (5 Part with Retics)- 05 Nos.
GeM bid reference No. : GEM/2019/B/473912 Dt. 04-01-2020
Bid End Date & Time : 24-01-2020 at 3 PM

Specification	Specification Name	Bid Requirement as per GeM	Additional conditions / Bid Requirement as per OSMCL
Standards	Certifications for the product	NIL	The quoted model should be USFDA and CE approved. CE certificate must be compliant with in-vitro diagnostic device (IVD) standard.

2. **Name of the item & Qty** : Automated Hematology Analyser (5 Part without Retics)- 08 Nos.
GeM bid reference No. : GEM/2019/B/ 487513 Dt. 04-01-2020
Bid End Date & Time : 24-01-2020 at 3 PM

Specification	Specification Name	Bid Requirement as per GeM	Additional conditions / Bid Requirement as per OSMCL
Standards	Certifications for the product	NIL	The quoted model should be USFDA and CE approved. CE certificate must be compliant with in-vitro diagnostic device (IVD) standard.

3. **Name of the item & Qty** : Fully Automatic Bio Chemistry Analyzer with Throughput (without ISE: 600 or more) – 09 Nos.
GeM bid reference No. : GEM/2019/B/ 487520 Dt. 04-01-2020
Bid End Date & Time : 24-01-2020 at 3 PM

Specification	Specification Name	Bid Requirement as per GeM	Additional conditions / Bid Requirement as per OSMCL
Standards	Certifications for the product	NIL	The quoted model should be USFDA and CE approved. CE certificate must be compliant with in-vitro diagnostic device (IVD) standard.

4. **Name of the item & Qty** : Fully Automatic Bio Chemistry Analyzer with Throughput (without ISE: 150 or more) – 33 Nos.
GeM bid reference No. : GEM/2019/B/ 487508 Dt. 04-01-2020
Bid End Date & Time : 24-01-2020 at 3 PM

Specification	Specification Name	Bid Requirement as per GeM	Additional conditions / Bid Requirement as per OSMCL
Standards	Certifications for the product	NIL	The quoted model should be USFDA and CE approved. CE certificate must be compliant with in-vitro diagnostic device (IVD) standard.

5. **Name of the item & Qty** : High Performance Liquid Chromatography (HPLC) System - 05 Nos.
GeM bid reference No. : GEM/2019/B/488534 Dt. 04-01-2020
Bid End Date & Time : 24-01-2020 at 3 PM

Specification	Specification Name	Bid Requirement as per GeM	Additional conditions / Bid Requirement as per OSMCL
Certification	Confirmity to quality management standards	NIL	Should be ISO 13485 certified for quality standards.

6. **Name of the item & Qty** : Horizontal Electrophoresis system - 13 Nos.
GeM bid reference No. : GEM/2019/B/487550 Dt. 04-01-2020
Bid End Date & Time : 24-01-2020 at 3 PM

Specification	Specification Name	Bid Requirement as per GeM	Additional conditions / Bid Requirement as per OSMCL
Standards	Certification for product	CE	The quoted model should be USFDA or CE approved.

7. **Name of the item & Qty** : Medical Centrifuge - 70 Nos.
GeM bid reference No. : GEM/2019/B/488456 Dt. 04-01-2020
Bid End Date & Time : 24-01-2020 at 3 PM

Specification	Specification Name	Bid Requirement as per GeM	Additional conditions / Bid Requirement as per OSMCL
Reports	ISO certification for Manufacturer	Yes / No	Yes, Should be ISO 13485 certified for quality standards.

8. **Name of the item & Qty** : Fume hoods for laboratory with cupboard / Grossing Station –05 Nos.
GeM bid reference No. : GEM/2019/B/488468 Dt. 04-01-2020
Bid End Date & Time : 24-01-2020 at 3 PM

Specification	Specification Name	Bid Requirement as per GeM	Additional conditions / Bid Requirement as per OSMCL
Standards / Certification	Certifications for the product	NIL	The quoted model should be USFDA or CE approved.
Reports	ISO certification for Manufacturer	NIL	Should be ISO 13485 certified for quality standards.

Note:

- 1) The prospective bidders must upload their Valid Certificates and product data sheet/technical brochure/user manuals along with the uploaded technical bid as mentioned in this notice as well as in the bid document of GeM for the purpose of technical evaluation.
- 2) In case bidders have not yet registered themselves for the above items in the GeM portal, they may register themselves at the earliest.

Sd/-
Managing Director, OSMCL