



NOTIFICATION NO.: 216 /OSMC/2021

Dt. 12-01-2021

PRE-BID AMENDMENT

Subject: Clarifications / amendment based on the queries received from prospective bidders against against the short tender for procurement of Cold Chain Equipment required for COVID-19 logistics & vaccination.

Short tender notification No.: 48 / OSMC/2021 Dtd. 05-01-2021

Total No. of queries received through e-mail and/or letters submitted at OSMC: **03 (Three) Nos.**
The queries raised by the prospective bidders against the above-mentioned tender reference were thread barely discussed and the following recommendations were made:

Sl.	Queries of the prospective bidders	Provision as per published bid / Original Technical Specification	Clarifications /Amendments of specification in response
TECHNICAL SPECIFICATIONS			
Item Name: Ice Lined Refrigerator (Big & Small)			
1.	Net Vaccine Storage Capacity of large ILR should be between 180-250 litre and small ILR : 80-130 litre instead of 50 litre (Small) & 100 Litre (Big)	Vaccine Storage Capacity: 50 litre (Small) & 100 Litre (Big) with tolerance of $\pm 10\%$.	No Change
2.	Ambient Temperature : should be +10 to +43 deg. C as per Indian Conditions instead of +5 to +45 deg. C.	Temperature Range: Temperature of a full vaccines to shall remain from +2 deg C to +8 deg C during continuous availability of energy at ambient temperature +5 to +45 deg. C.	Amended as: Temperature Range: Temperature of full vaccines shall remain from +2 deg C to +8 deg C during continuous availability of power supply at ambient temperature of +10 to +43 deg. C.
3.	Hold Over Time : Minimum 40 hours or more from 2 deg.C To 8 deg.C at ambient temperature +10 to +43 deg. C as per Indian Conditions instead of +5 to +45 deg. C ambient temperature	Hold Over Time : Minimum 40 hours or more from +2 deg C to +8 Deg C at +5 to +45 deg. C ambient temperature	Amended as: Hold Over Time : Minimum 40 hours or more to maintain a temperature range of +2 deg C to +8 Deg C at +10 to +43 deg. C ambient temperature.
4.	Defrost facility : Automatic or manual defrosting. In case of Automatic defrosting, provision for defrosting program should be available whereas as per MoHFW Anti-	Defrost facility: Automatic or manual defrosting. In case of Automatic defrosting, provision for defrosting program should be available.	Amended as: Defrost facility: Automatic or manual defrosting or should have Anti-Freeze Grade A feature.



ODISHA STATE MEDICAL CORPORATION LTD.
(A Govt. of Odisha Undertaking)
(ASSURING QUALITY, SAVING LIVES)

Phone No. : (0674) 2380608,
 Website: www.osmcl.nic.in,
 E-mail: proc.osmcl.od@nic.in

	Freeze Grade A mentioned.		
5.	Tray Type : Powder Coated Wire Shelves should also be allowed along with Adjustable upright SS mess type.	Tray Type: Adjustable upright SS mess type	Amended as: Tray Type: Stainless Steel (SS) OR Powder Coated Wire Shelves / stackable buckets.
6.	Coolant : Ethylene Glycol of high Specific heat capacity is not applicable for Vertical Ice Lined Refrigerator.	Coolant : Ethylene Glycol of high Specific heat capacity	Amended as: Coolant (if required): Ethylene Glycol (or similar) having high Specific heat capacity to maintain the required hold over time as defined in the technical specification.
7.	Safety : (i) Resettable over current breaker shall be fitted before stabilizer for protection. (ii) Fuses for Compressor and Temperature controller : Whereas we are providing inbuilt voltage stabilizer.	Safety : (i) Resettable over current breaker shall be fitted before stabilizer for protection. (ii) Fuses for Compressor and Temperature controller.	Amended as: Safety : (i) Resettable over current breaker shall be fitted before external stabilizer or internal stabilizer for protection. (ii) Fuses for Compressor and Temperature controller. (Point Deleted)
		Servo controlled Stabilizer : Appropriate rating of automatic servo-controlled stabilizer ISI marked (Copper core) 110V to 270V with High/Low Voltage cut-off and time-delay restart must be provided with the machine.	Amended as: External or internal stabilizer: (i) In case of external stabilizer: Appropriate rating of stabilizer ISI marked (Copper core) 110V to 270V with High/Low Voltage cut-off and time-delay restart must be provided with the machine and Should have appropriate facility for voltage & current over load protection (ii) In case of internal stabilizer: Should stabilize the voltage within the input range 110V to 270V with appropriate facility for voltage and current overload protection. N.B: Servo controlled Stabilizer is not required and is deleted.



ODISHA STATE MEDICAL CORPORATION LTD.
(A Govt. of Odisha Undertaking)
(ASSURING QUALITY, SAVING LIVES)

Phone No. : (0674) 2380608,
Website: www.osmcl.nic.in,
E-mail: proc.osmcl.od@nic.in

8.	Request to remove WHO approved document	Quality Standard Certifications for	No Change
9.	WHO certificate under WHO/PQS/E03/RF03.4 is applied and approval is awaited. Request to grant us to participate.	Product: The quoted product should meet the Standard of WHO/PQS/E03/RF03.1 for Ice Lined Refrigerators. (model approved document should be furnished)	
Item Name: DEEP FREEZER (LARGE)			
1.	Audio/Visual Alarm for Temperature High & Low so as to help maximum participation	Audio Visual Alarms & indications for Temperature High & Low, Door Open, power failure.	Amended as: Audio/Visual alarms or indicators for Temperature High & Low and power failure. Door should have locking facility to ensure proper closing and to avoid accidental or unintentional door opening.
TECHNO-COMMERCIAL PARAMETERS:			
1.	MAKE IN INDIA : As per The Ministry of Commerce and Industry Department of Industrial policy had vide their order No. P-45021/2/2017- B.E.-II with revision dated 04.06.2020, purchase preference to Local supplier under "Make In India" to be provided.	Additional Clause	No Change

N.B:

1. Amendment to dates:

The last of submission has been extended and revised dates are as follows:

- **Amended Last date for bid submission : 20.01.2021, 04:00 PM**
 - **Amended Last date technical bid opening : 20.01.2021, 04:30 PM**
2. The amendments mentioned above are to be treated as amendments in the general term(s) and condition(s) and scope of work of the above tender reference. All other terms conditions remain unchanged.
3. The amendments mentioned above are to be treated as amendments pertaining to price bid (if applicable).

Sd/-
General Manager, Operation (Equipment)
Odisha State Medical Corporation