



Odisha State Medical Corporation Limited (OSMCL)

(Assuring Quality, Saving Lives)

In front of Ram Mandir, Convent Square, Unit – III,
Bhubaneswar -751 001

Tel. : (0674) 2380660/2380608/2380950

Website : www.osmcl.nic.in , Email : logistics.osmcl.od@nic.in

Notice No.- 3364

Date:- 24.03.2020

PUBLIC NOTICE

OSMCL is going to procure following items urgently. Interested parties may submit their willingness to supply the following items:

Sl. No.	Name of Item with specification	Approved Price (including GST)	Contact Number
1.	<i>Personal Protection Equipment (PPE) Kit</i>	<i>Rs.1087.47</i>	<i>8895951226</i>
2.	<i>Triple Layer Face Mask</i>	<i>Rs.10.00</i>	<i>9438183574</i>
3.	<i>N 95 Disposable Particulate Respirator</i>	<i>Rs.49.61</i>	<i>9439315941</i>

INCENTIVES

Incentives will be paid over and above the approved price in the following manners

1. @ 50 % of the delivered value- if delivered within 7 days of the order date.
2. @ 25 % of the delivered value- if delivered within 15 days of the order date.
3. @ 10 % of the delivered value- if delivered within 30 days of the order date.

The interested suppliers are requested to submit their willingness alongwith the quantity to the office of “Managing Director, Odisha State Medical Corporation Ltd. In front of Ram Mandir, Convent Square, Unit – III, Bhubaneswar -751 001” in the e-mail Id logistics.osmcl.od@nic.in for subsequent issue of supply order. For detailed specification of the above items, please visit our official website.

Sd/-
Managing Director
OSMCL

Technical Specification for PPE (Personnel Protection Equipment)

1. PPE Kit

1.1 Gloves (Double pair)

- Nitrile
- Non-sterile
- Powder free
- Outer gloves preferably reach mid –forearm (minimum 280 mm total length)
- Different sizes (6.5 & 7)
- Quality compliant with the below standards, or equivalent:
 - a) EU standard directive 93/42/EEC Class I, EN 455
 - b) EU standard directive 89 / 686/EEC Category III, EN 374
 - c) ANSI/SEA 105-2011
 - d) ASTM D6319-10

1.2 Coverall (medium and large)

- Impermeable to blood and body
- Single use
- Avoid culturally unacceptable color e.g. black
- Light colors are preferable to better detect possible contamination
- Thumb/finger loops to anchor sleeves in place
- Quality compliant with following standard
 - a. Meets or exceeds to ISO 16603 class 3 exposure pressure, or equivalent

1.3 Goggles

- With transparent glasses, zero power, well fitting , covered from all sides with elastic band / or adjustable holder
- Good seal with the skin of the face
- Flexible frame to easily fit all face contours without too much pressure
- Covers the eyes and the surrounding areas the accommodates for prescription glasses
- Fog and scratch resistant
- Adjustable band to secure firmly so as not to become loose during clinical activity
- Indirect venting to reduce fogging
- May be re-usable (provided appropriate arrangements for decontamination are in place or disposable
- Quality compliant with the below standards, or equivalent :
 - a. EU standard directive 86/686/EEC, EN 166/2002
 - b. ANSI/SEA Z87.1-2010

1.4 N-95 Masks

- Shape that will not collapse easily
- High filtration efficiency
- Good breathability, with expiratory valve
- Quality complain with standards for surgical N95 respirator
 - a. NIOSH N95, EN 149 FFP2, or equivalent
- Fluid resistance : minimum 80 mmHg pressure based on ASTM F1862, ISO 22609, or equivalent
- Quality complaint with standards for particulate respirator that can be worn with full face shield

2. Face Shield

- ▣ Made of clear plastic and provides good visibility to both the wearer and the patient
- ▣ Adjustable band to attach firmly around the head and fit snugly against the forehead
- ▣ Fog resistant (preferable)
- ▣ Completely covers the sides and length of the face
- ▣ May be re-usable (made of material which can be cleaned and disinfected) or disposable
- ▣ Quality compliant with the below standards , or equivalent :
 - a. EU standard directive 86/686/EEC, EN 166/2002
 - b. ANSI/SEA Z87.1-2010.

3. Triple Layer Surgical Mask

- ▣ Three layered surgical mask of non-woven material with nose piece, having filter efficiency of 99% for 3 micron particle size.
- a. ISI specification or equivalent.

4. Gloves

- ▣ Nitrile
- ▣ Non-sterile
- ▣ Powder free
- ▣ Outer gloves preferably reach mid-forearm (minimum 280mm total length)
- ▣ Different sizes (6.5 & 7)
- ▣ Quality compliant with the below standards, or equivalent :
 - a. EU standard directive 93/42/EEC Class I, EN 455
 - b. EU standard directive 89 /686/EEC Category III, EN 374
 - c. ANSI/SEA 105-2011
 - d. ASTM D6319-10

All items to be supplied need to be accompanied with certificate of analysis from national /international organizations/ labs indicating conformity to standards

All items : Expiry 5 years

Specification of Triple (three) Layer Face Mask

- ▣ (9cm x 17cm approx.),
- ▣ Nose clip adaptable.
- ▣ Tie on mask of nonwoven, ▣ hypoallergenic
- ▣ 3 ply construction with filter in between offering >99% standard with 4 tie strings.)

Specification of N 95 Disposable Particulate respirator

- ▣ Shape that will not collapse easily
- ▣ High filtration efficiency
- ▣ Good breathability, with expiratory valve
- ▣ Quality complaint with standards for surgical N95 respirator NIOSH N95, EN 149 FFP2, or equivalent
- ▣ Fluid resistance: minimum 80 mmHg pressure based on ASTM F1862, ISO 22609, or Equivalent
- ▣ Quality complaint with standards for particulate respirator that can be worn with full Face shield

**Sd/-
Managing Director
OSMCL**