



Notification No.: 843

Date 31.01.2025

BIDDERS' SELECTION / REJECTION LIST

Tender Reference No.: OSMCL/2023-24/ Furniture/03

- ❖ TENDER NAME: e-Tender – (FOR SUPPLY, INSTALLATION & COMMISSIONING OF FURNITURE ON RATE CONTRACT BASIS)
- ❖ ITEM NAME: WHEELCHAIR FOLDING

I. LIST OF BIDDERS WHICH ARE TECHNICALLY RESPONSIVE AND QUALIFIED FOR ON-LINE PRICE BID OPENING

Sl.	Name of the Qualifying Bidders
1	M/s. Mediplus

II. LIST OF BIDDERS & THEIR OFFERED ITEMS WHICH ARE TECHNICALLY NON RESPONSIVE AND NOT QUALIFIED

Sl	Name of Bidder	Reason of Rejection
1.	M/s. HHW CARE PRODUCTS INDIA PVT LTD	1) Bidder has not submitted the ISO 45001 certificate, as Per Bid Requirement. 2) Bidder has not submitted SS 304 grade certificate from any Govt. or NABL approved lab accredited as per ISO 17025, as Per Bid Requirement. 3) Bidder has not submitted the impact test, bend test, salt spray chamber test reports from any Govt. or NABL approved lab accredited as per ISO 17025 for quoted item, as Per Bid Requirement. 4) Bidder has not submitted the CDSCO (License /online registration acknowledgment) certificate as Per Bid Requirement. 5) The submitted ISO 13485 certificate is not issued from a notified body or ICMED 13485 (with or without plus) certificate issued from certification bodies accredited by NABCB or issued from certification bodies accredited by NABCB/Nationally Recognized Accreditation Board under IAF MLA., as Per Bid Requirement.
2.	M/s. HOSCO PRIVATE LIMITED	1) The quoted model i.e HPL-59 of the Wheel chair is not in the submitted CE certificate 2) Bidder has not submitted ISO 14001 certificate, as Per Bid Requirement. 3) Bidder has not submitted SS 304 grade certificate from any Govt. or NABL approved lab accredited as per ISO 17025, as Per Bid Requirement. 4) Bidder has not submitted the impact test, bend test, salt spray chamber test reports from any Govt. or NABL approved lab accredited as per ISO 17025, as Per Bid Requirement.

3.	M/s. Medico International	<p>1) The quoted model i: e AS-170 is not mentioned in the submitted CE certificate.</p> <p>2) Bidder has not submitted SS 304 grade certificate from any Govt. or NABL approved lab accredited as per ISO 17025, as Per Bid Requirement.</p> <p>3) Bidder has not submitted the impact test, bend test, salt spray chamber test reports from any Govt. or NABL approved lab accredited as per ISO 17025, as Per Bid Requirement.</p> <p>4) Bidder has not submitted the product data sheet / brochure for the quoted model. Hence the technical specification could not be verified.</p> <p>5) Instead of polished stainless steel construct, bidder has submitted mild steel epoxy powder coated as per bid requirement.</p> <p>6) Instead of Nylon Handgrips, bidder has submitted PVC handgrip as per bid requirement.</p>
4.	M/s. Midmark India Pvt.Ltd	<p>1) The quoted model i:e 1039 (Privy Folding wheel Chair) of the Wheel Chair Folding is not in the submitted CE certificate</p> <p>2) Bidder has not submitted SS 304 grade certificate from any Govt. or NABL approved lab accredited as per ISO 17025 as per bid requirement.</p> <p>3) Bidder has not submitted the impact test, bend test, salt spray chamber test reports from any Govt. or NABL approved lab accredited as per ISO 17025, as Per Bid Requirement.</p> <p>4) Instead of Aluminium leg rest, bidder has submitted plastic foot plate as per bid requirement.</p>
5.	M/s. Nilkamal Limited	<p>1) Bidder has not submitted SS 304 grade certificate from any Govt. or NABL approved lab accredited as per ISO 17025 as Per Bid Requirement.</p> <p>2) Bidder has submitted semi fowler bed with side rail in format T10 (STATEMENT OF DEVIATION) and submitted wheelchair folding (PARAWISE COMPLIANCE TO TECHNICAL SPECIFICATION OF THE PRODUCT) in format T11.</p> <p>3) Bidder has not submitted the parawise / point wise compliance to technical specification in format T11, Hence the technical parameter could not be verified.</p>
6.	M/s. Ostrich Mobility Instruments Pvt Ltd	<p>1) The quoted model i: e T130SSE of the Wheel Chair Folding is not in the submitted CE certificate.</p> <p>2) Bidder has not submitted SS 304 grade certificate from any Govt. or NABL approved lab accredited as per ISO 17025 as per bid requirement.</p> <p>3) Bidder has not submitted the impact test, bend test, salt spray chamber test reports from any Govt. or NABL approved lab accredited as per ISO 17025, as Per Bid Requirement.</p> <p>4) Bidder has not submitted the CDSCO (License /online registration acknowledgment).</p>
7.	M/s. Riya Fibro Tech	<p>1) Bidder has not submitted the product quality standard certificate i: e USFDA/CE/BIS for the quoted model as per bid requirement.</p> <p>2) Bidder has not submitted SS 304 grade certificate from any Govt. or NABL approved lab accredited as per ISO 17025 as per bid requirement.</p> <p>3) Bidder has not submitted the impact test, bend test, salt spray chamber test reports from any Govt. or NABL approved lab accredited as per ISO 17025 for quoted item, as Per Bid Requirement.</p> <p>4) Bidder has not submitted the CDSCO (License /online registration acknowledgment) for the quoted product as Per Bid Requirement.</p> <p>5) Bidder has not submitted the product data sheet / brochure for</p>

		<p>the quoted model I: eVWC1. Hence the technical specification could not be verified.</p> <p>6) The turnover of OEM as mentioned in Format-T8 could not be verified from the supporting Annual audited statements.</p> <p>7) The turnover of Distributor for the FY 2020-21 & 2022-23 as mentioned in Format-T8 could not be verified from the supporting Annual audited statements.</p> <p>8) Not Submitted the Proof of Past supply (PO copies) of Distributor.</p>
8.	M/s. Vedgarbh Enterprises Pvt.Ltd.	1) Bidder has not submitted the bid document cost.

N.B.: Since the tender resulted in single responsive bid, the Committee further recommended for floating of retender for the same.



GM-Operation (Equipment)
Odisha State Medical Corporation

