



Notification No.: 6964

Date 19.10.2024

BIDDERS' SELECTION / REJECTION LIST

Tender Reference No. : OSMCL/2023-24/ Radiology_Equipment/06

- ❖ TENDER NAME: e-Tender – (Supply, Installation & Commissioning of Radiology Equipments on Rate Contract Basis)
- ❖ ITEM NAME: CR Scanner with 3 Tray

I. **LIST OF BIDDERS & THEIR OFFERED ITEMS WHICH ARE NON RESPONSIVE AND NOT QUALIFIED**

| SI | Name of Bidder | Reason of Rejection |
|----|-----------------------------------|---|
| 1. | M/s. JK Medical System | <ol style="list-style-type: none">1. The quoted model digitizer does not have 20 pixel/mm (minimum) for high resolution cassette reading as per the submitted Product Datasheet.2. The Bidder has not submitted Land Border sharing Declaration.3. The Committee observed that the OEM, M/s. CARESTREAM Health India Pvt Ltd has authorised M/s. EDUSOFT HEALTHCARE PVT. Ltd. as well as M/S JK MEDICAL SYSTEMS for the same item i.e. CR System (3 Tray). The committee recommended to reject both the bidders i.e. M/s. EDUSOFT HEALTHCARE Pvt. Ltd. & M/s. JK MEDICAL SYSTEMS as per tender clause No. 5.2.7. |
| 2. | M/s. Edusoft Healthcare Pvt. Ltd. | <ol style="list-style-type: none">1. Has submitted EN ISO 13485 certificate for the manufacturer Rayco (Shanghai) Medical Products Company Limited instead of the quoted product manufacturer i.e Carestream Health India Pvt. Ltd as per the submitted format T-2.2. Has submitted the CDSCO online registration screenshot details of Radiographic film instead of the quoted medical device.3. The quoted model digitizer does not have 20 pixel/mm (minimum) for high resolution cassette reading as per the submitted Product Datasheet.4. The quoted model digitizer has minimum image preview time more than 60 sec as per the submitted Product data |

| | | |
|---|-----------------------------|---|
| | | <p>sheet.</p> <ol style="list-style-type: none"> 5. Bidder has not submitted Compliance in required format for provision of Online UPS of suitable rating for back up for 30 minutes power backup for the whole system. 6. Bidder has not submitted Compliance in required format regarding Cassette Reader, Printer and Console (acquisition & processing) software should be from Parent manufacturer/Same OEM. 7. The Bidder has not submitted Land Border sharing Declaration. 8. In Format-T4, Bidder declared as OEM, however submitted Format T7 from OEM, M/s. CARESTREAM Health India Pvt. Ltd. The Committee observed that the OEM, M/s. CARESTREAM Health India Pvt. Ltd has authorised M/s. EDUSOFT HEALTHCARE PVT. Ltd as well as M/S JK MEDICAL SYSTEMS for the same item i.e. CR System (3 Tray). The committee recommended to reject both the bidders i.e. M/s. EDUSOFT HEALTHCARE Pvt. Ltd. & M/s. JK MEDICAL SYSTEMS as per tender clause No. 5.2.7. 9. There is mismatch of Turn over figures between Audited Account Statement & Format-T8 for Financial year 2020-21 of the Bidder. 10. Format-T8 submitted but no Supporting balance sheet & Audited Account Statement of OEM submitted |
| 3 | M/s. AXESS MEDICOM PVT.LTD. | <ol style="list-style-type: none"> 1. The Bidder has not submitted Land Border sharing Declaration. 2. In T7 Format, the Bidder has declared the name of Factory as China, which is prohibited for participation due to imposition of restriction on Bidder from a country which share a land border with India based on the Notification No. F.7/10/2021-PPD (1) Order (Public Procurement No. 4) regarding Restrictions under Rule 144 (xi) of the General Financial Rule (GFRs), 2017 & Notification No. 27945/F Dated. 16.10.2020 of Govt. of Odisha, (Finance Department) |

❖ **N.B:** Since the tender resulted in non-responsive bid, the Committee further recommended for cancellation of the tender for the item CR Scanner with 3 Tray & floating of retender for the same.


19.10.24

GM-Operation (Equipment)
Odisha State Medical Corporation