



ODISHA STATE MEDICAL CORPORATION LTD.
(A Govt. of Odisha Undertaking)
(Assuring Quality, Saving Lives)

Phone No. : (0674) 2380608,
Website: www.osmcl.nic.in,
E-mail: adm.osmcl@nic.in

Notification No. 704

Date: 28.01.2025

NOTICE

Sub : OSMCL –E- Tender for SUPPLY OF SURGICAL, SUTURES, SYRINGES AND OTHER ITEMS (RE-TENDER) for the year 2024-25– Technical Bid Evaluation –status– Published – regarding.

Bid Ref. No. OSMCL/2024-25/(SUT.-SURG.)-DHS- RE-TENDER/09 dtd.29/11/2024

Technical bids submitted in response to the e-tender referred above for **SUPPLY OF SURGICAL, SUTURES, SYRINGES AND OTHER ITEMS (RE-TENDER) for the year 2024-25** were opened on 03.01.2025 and scrutinized. The **provisional** list of eligible and rejected bidders with their items after technical bid evaluation is published at Annex – I. Representations, if any, on this notice should be submitted to the Tender Inviting Authority within 3 days from the publication of this notice i.e on or before **Dt: 30.01.2025, 5.00PM** through e-mail ([drugs-osmcl@gov.in.](mailto:drugs-osmcl@gov.in)), after that no representation from any bidder will be taken for consideration.

N.B.: No new documents will be entertained /taken for consideration. Only the Bidder has to mention the **page No.** if already uploaded the document in the e-tender portal regarding the representation.

Encl: Annex – I


GM Operations (D & S)
OSMC Ltd., Odisha

**Provisional list of Eligible and Rejected Bidders with their items
for Surgical, Suture & Other Items(Re-Tender) 2024 - 25**

Bid Reference No. : OSMCL/2024-25/(SUT.-SURG.)-DHS- RE-TENDER/09

Sl No.	Name of the Item	Specification	Reason for Rejection	Remarks
1	1	M/s. MEDISAFE INTERNATIONAL ,Delhi (North Delhi)		
1	S02331 Nebuliser set with T piece	Nebuliser-1 No., "T" Piece with 15mm Female for endotracheal tube and 22mm Female fitting for the mouth piece -1 No., flexible medical grade 2mtr long multichannel tube, kink resistant - 1 No.		Qualified.
2	S02343 Resusciator Kit Manual (Adult)	Ambu bag with MDI port, Star lumen kink resistant oxygen tubing, reservoir bag, pressure limiting valve and silicone mask, made up from 100% silicone, bag re-expansion property 360 degree swivel joint. Autoclavable, connectors (15/22 mm). Size - Adul		Qualified.
3	S02344 Resusciator Kit Manual (Child)	Ambu bag with MDI port. Star lumen kink resistant oxygen tubing, reservoir bag, pressure limiting valve and silicone mask, made up from 100% silicone, bag re-expansion property 360 degree swivel joint. Autoclavable, connectors (15/22 mm). Size - Chil		Qualified.
4	S02345 Resusciator Kit Manual (Neonate)	Ambu bag with MDI port. Star lumen kink resistant oxygen tubing, reservoir bag, pressure limiting valve and silicone mask, made up from 100% silicone, bag re-expansion property 360 degree swivel joint. Autoclavable, connectors (15/22 mm). Size - Infa		Qualified.
5	S02432 Bain Circuit	Sterile pack, non-toxic, pyrogen free, silicon product; sterilisable (autoclave / ETO)	Drug endorsement for Non Sterile instead of Sterile and material is not silicon.	Not Qualified.
6	S02451 Foleys Catheter, size 8 FG	3way, Sterile pack, non-toxic, natural rubber latex, pyrogen free, size 8 FG,with CE certification.		Qualified.
7	S02452 Foleys Catheter, size 10 FG	3way, Sterile pack, non-toxic,natural rubber latex, pyrogen free, size 10 FG,with CE certification.		Qualified.
8	S02453 Foleys Catheter, size 12 FG	3way, Sterile pack, non-toxic,natural rubber latex, pyrogen free, size 12 FG,with CE certification.		Qualified.
9	S02471 Urometer	2000 ml capacity, graduated, sterile, soft and kink resistant tubing, clear and transparent chamber.	Drug endorsement for Urine bag instead of Urometer.	Not Qualified.

Sl No.	Name of the Item	Specification	Reason for Rejection	Remarks
10	S02476 Vaginal Speculum	Each Sterile disposable, Small		Qualified.
11	S02477 Vaginal Speculum	Each Sterile disposable, Medium		Qualified.
12	S02478 Vaginal Speculum	Each Sterile disposable, Large		Qualified.

2 2 M/s. PLASTI SURGE INDUSTRIES PVT. LTD. ,Maharashtra (Amravati)

1	S02029 Disposable Apron	Long body, sterile, disposable, Not less than 25 GSM material, Colour-Green/Blue IS 4501: 198(R2006) and ASTM F2407.	No BIS & ASTM F2407 certificate as per clause no. 5.2.12	Not Qualified.
2	S02042 Disposable PPE Kit	1. Cap. 1No. 2. Gown 1No. 3. Goggles 1No. 4. Face Mask 1No. 5. Sterile Gloves 2Pairs. 6. Shoe Cover. All the Items must be packed in a single polypack.		Qualified.
3	S02055 Disposable Surgical Head Cap	Each disposable		Qualified.
4	S02079 Triple (three) layer face mask	(9cm x 17cm approx.), nose clip adaptable.(Tie on mask of no-woven, hypoallergenic 3 ply construction with filter in ply construction with filter in between offering >99% standard with Ear strings.) with IS 16289:2014.		Qualified.
5	S02118 Disposable Delivery Kit (ETO Sterilised)	AS PER TENDER SPECIFICATION	Components mentioned in MD Lic not matching with specification.	Not Qualified.
6	S02283 Drape Sheet Sterilised	Sterilised; Size: 100cm x 150cm, Plastic	Drape Sheet size is not as per specification.	Not Qualified.
7	S02461 Kelley s Pad	Disposable, non-woven, 45GSM, Large	Drug endorsement for under buttock drape instead of Kelley s Pad.	Not Qualified.
8	S02470 Tourniquet	Each with heavy duty zipper	No Drug endorsement.	Not Qualified.

3 3 M/s. ASHRI MEDICO SURGICAL PVT. LTD. ,MP (Indore)

1	S02029 Disposable Apron	Long body, sterile, disposable, Not less than 25 GSM material, Colour-Green/Blue IS 4501: 198(R2006) and ASTM F2407.	No BIS & ASTM F2407 certificate as per clause no. 5.2.12	Not Qualified.
2	S02118 Disposable Delivery Kit (ETO Sterilised)	AS PER TENDER SPECIFICATION		Qualified.
3	S02471 Urometer	2000 ml capacity, graduated, sterile, soft and kink resistant tubing, clear and transparent chamber.	Less Turnover of 5 Crores for the year 2021 - 22.	Not Qualified.

Sl No.	Name of the Item	Specification	Reason for Rejection	Remarks
--------	------------------	---------------	----------------------	---------

4 4 M/s. INNOVATIVE LINEN CO. PVT. LTD. ,Indore

1	S02029 Disposable Apron	Long body, sterile, disposable, Not less than 25 GSM material, Colour-Green/Blue IS 4501: 198(R2006) and ASTM F2407.	No (BIS & ASTM F2407), Non Conviction & T8 certificate submitted for non sterile as per clause no. 5.2.12., 5.2.10 & 5.2.9 respectively.	Not Qualified.
2	S02042 Disposable PPE Kit	1. Cap. 1No. 2. Gown 1No. 3. Goggles 1No. 4. Face Mask 1No. 5. Sterile Gloves 2Pairs. 6. Shoe Cover. All the Items must be packed in a single polypack.	No Drug endorsement.	Not Qualified.
3	S02055 Disposable Surgical Head Cap	Each disposable	No Non Conviction & T8 certificate submitted as per clause no. 5.2.10 & 5.2.9 respectively.	Not Qualified.
4	S02078 N 95 Disposable Particulate respirator	a) Shape that will not collapse easily b) High filtration efficiency c) Good breathability, with expiratory valve d) Quality compliance with standards for surgical N95 respirator(i) NIOSH N95, EN 149 FFP2, or equivalent e) Fluid resistance : minimum	No Non Conviction & T8 certificate submitted as per clause no. 5.2.10 & 5.2.9 respectively.	Not Qualified.
5	S02079 Triple (three) layer face mask	(9cm x 17cm approx.), nose clip adaptable.(Tie on mask of no-woven, hypoallergenic 3 ply construction with filter in ply construction with filter in between offering >99% standard with Ear strings.) with IS 16289:2014.	No BIS, No Non Conviction & T8 certificate submitted against MD Lic. No 000613 as per clause no. 5.2.12, 5.2.10 & 5.2.9 respectively.	Not Qualified.
6	S02118 Disposable Delivery Kit (ETO Sterilised)	AS PER TENDER SPECIFICATION	Non Conviction & T8 certificate submitted for non sterile as per clause no. 5.2.10 & 5.2.9 respectively.	Not Qualified.
7	S02283 Drape Sheet Sterilised	Sterilised; Size: 100cm x 150cm, Plastic	No Drug endorsement.	Not Qualified.
8	S02442 Disposable C-Arm Cover	Sterile, 80cm x 90cm, each pair with adhesive sticker, with CE certification	No Drug endorsement.	Not Qualified.
9	S02461 Kelley s Pad	Disposable, non-woven, 45GSM, Large		Qualified.

5 5 M/s. MERIL ENDO SURGERY PVT. LTD. ,Maharashtra (Mumbai)

1	S01183 Polyglactin 7-0	Size:- 7-0, Absorbable surgical suture USP, 0.5 metric, 30cm (+/- 5-10%), spatula, double lumen curved (A7102), with CE certification	Invalid CE certificate as per clause No. 5.2.12.	Not Qualified.
---	------------------------	---	--	----------------

SI No.	Name of the Item	Specification	Reason for Rejection	Remarks
--------	------------------	---------------	----------------------	---------

6 6 M/s. HEALTHCARE MEDICAL DEVICES PVT. LTD. ,Indore

1	D31042 ECG / US Jelly	5Kg / Jar		Qualified.
2	S02055 Disposable Surgical Head Cap	Each disposable		Qualified.
3	S02079 Triple (three) layer face mask	(9cm x 17cm approx.), nose clip adaptable.(Tie on mask of no-woven, hypoallergenic 3 ply construction with filter in ply construction with filter in between offering >99% standard with Ear strings.) with IS 16289:2014.	Less Turnover of 5 Crores for the year 2021 - 22.	Not Qualified.
4	S02118 Disposable Delivery Kit (ETO Sterilised)	AS PER TENDER SPECIFICATION	Safe del. Kit mentioned instead of Disp. del. kit in Drug endorsement Non Conviction & T8 certificate as per clause no. 5.2.10 & 5.2.9 respectively.	Not Qualified.
5	S02283 Drape Sheet Sterilised	Sterilised; Size: 100cm x 150cm, Plastic	Less Turnover of 5 Crores for the year 2021 - 22.	Not Qualified.
6	S02461 Kelley s Pad	Disposable, non-woven, 45GSM, Large	Less Turnover of 5 Crores for the year 2021 - 22.	Not Qualified.

7 7 M/s. PRIMEWEAR HYGINE (INDIA) PRODUCTS LTD. ,Maharashtra (Palghar)

1	S02029 Disposable Apron	Long body, sterile, disposable, Not less than 25 GSM material, Colour-Green/Blue IS 4501: 198(R2006) and ASTM F2407.	No BIS & ASTM F2407 certificate as per clause no. 5.2.12.	Not Qualified.
2	S02042 Disposable PPE Kit	1. Cap. 1No. 2. Gown 1No. 3. Goggles 1No. 4. Face Mask 1No. 5. Sterile Gloves 2Pairs. 6. Shoe Cover. All the Items must be packed in a single polypack.		Qualified.
3	S02118 Disposable Delivery Kit (ETO Sterilised)	AS PER TENDER SPECIFICATION		Qualified.
4	S02283 Drape Sheet Sterilised	Sterilised; Size: 100cm x 150cm, Plastic	Drape Sheet size is not as per specification.	Not Qualified.
5	S02442 Disposable C-Arm Cover	Sterile, 80cm x 90cm, each pair with adhesive sticker, with CE certification	Invalid CE certificate as per clause No. 5.2.12.	Not Qualified.
6	S02461 Kelley s Pad	Disposable, non-woven, 45GSM, Large		Qualified.

SI No.	Name of the Item	Specification	Reason for Rejection	Remarks
8	8	M/s. FLEXICARE MEDICAL (INDIA) PVT. LTD. ,Chennai		
1	S02343 Resusciator Kit Manual (Adult)	Ambu bag with MDI port. Star lumen kink resistant oxygen tubing, reservoir bag, pressure limiting valve and silicone mask, made up from 100% silicone, bag re-expansion property 360 degree swivel joint. Autoclavable, connectors (15/22 mm). Size - Adul		Qualified.
2	S02344 Resusciator Kit Manual (Child)	Ambu bag with MDI port. Star lumen kink resistant oxygen tubing, reservoir bag, pressure limiting valve and silicone mask, made up from 100% silicone, bag re-expansion property 360 degree swivel joint. Autoclavable, connectors (15/22 mm). Size - Chil		Qualified.
3	S02345 Resusciator Kit Manual (Neonate)	Ambu bag with MDI port. Star lumen kink resistant oxygen tubing, reservoir bag, pressure limiting valve and silicone mask, made up from 100% silicone, bag re-expansion property 360 degree swivel joint. Autoclavable, connectors (15/22 mm). Size - Infa		Qualified.

Sl No.	Name of the Item	Specification	Reason for Rejection	Remarks
--------	------------------	---------------	----------------------	---------

9 9 M/s. AMKAY PRODUCTS LTD. ,Maharashtra (Palghar)

1	S02029 Disposable Apron	Long body, sterile, disposable, Not less than 25 GSM material, Colour-Green/Blue IS 4501: 198(R2006) and ASTM F2407.	No (BIS & ASTM F2407), Drug endorsement & T8 certificate submitted for non sterile as per clause no. 5.2.12. & 5.2.9 respectively.	Not Qualified.
2	S02042 Disposable PPE Kit	1. Cap. 1No. 2. Gown 1No. 3. Goggles 1No. 4. Face Mask 1No. 5. Sterile Gloves 2Pairs. 6. Shoe Cover. All the Items must be packed in a single polypack.		Qualified.
3	S02055 Disposable Surgical Head Cap	Each disposable		Qualified.
4	S02079 Triple (three) layer face mask	(9cm x 17cm approx.), nose clip adaptable.(Tie on mask of no-woven, hypoallergenic 3 ply construction with filter in ply construction with filter in between offering >99% standard with Ear strings.) with IS 16289:2014.		Qualified.
5	S02283 Drape Sheet Sterilised	Sterilised; Size: 100cm x 150cm, Plastic		Qualified.
6	S02343 Resusciator Kit Manual (Adult)	Ambu bag with MDI port. Star lumen kink resistant oxygen tubing, reservoir bag, pressure limiting valve and silicone mask, made up from 100% silicone, bag re-expansion property 360 degree swivel joint. Autoclavable, connectors (15/22 mm). Size - Adul		Qualified.
7	S02344 Resusciator Kit Manual (Child)	Ambu bag with MDI port. Star lumen kink resistant oxygen tubing, reservoir bag, pressure limiting valve and silicone mask, made up from 100% silicone, bag re-expansion property 360 degree swivel joint. Autoclavable, connectors (15/22 mm). Size - Chil		Qualified.
8	S02345 Resusciator Kit Manual (Neonate)	Ambu bag with MDI port. Star lumen kink resistant oxygen tubing, reservoir bag, pressure limiting valve and silicone mask, made up from 100% silicone, bag re-expansion property 360 degree swivel joint. Autoclavable, connectors (15/22 mm). Size - Infa		Qualified.
9	S02461 Kelley s Pad	Disposable, non-woven, 45GSM, Large	No Drug endorsement.	Not Qualified.
10	S02471 Urometer	2000 ml capacity, graduated, sterile, soft and kink resistant tubing, clear and transparent chamber.	Drug endorsement for Urine bag instead of Urometer.	Not Qualified.

SI No.	Name of the Item	Specification	Reason for Rejection	Remarks
--------	------------------	---------------	----------------------	---------

10 10 M/s. INDIA MEDTRONIC PVT. LTD. ,Gurugram (Haryana)

1	S01126 Black Braided Silk 5-0	5-0, Curved Round Body, double armed sutures spatulated needles, with CE certification	No CE certificate & T8 certificate as per clause no. 5.2.12 & 5.2.9 respectively.	Not Qualified.
2	S01136 Polypropylene 7-0	Size-7-0, 6 / 5 mm (+/- 5-10%),length: 45cm (+/- 5-10%), 1/2 circle round bodied / Taper point, U.S.P, with CE certification	No CE certificate as per clause No. 5.2.12.	Not Qualified.
3	S01138 Polypropylene 8-0	Size-8-0, 9.3 mm,length: 45cm (+/- 5-10%), 1/2 circle round bodied / Taper point, U.S.P, with CE certification	No CE certificate as per clause No. 5.2.12.	Not Qualified.
4	S01190 Polyester 2	Size- 2, braided, 75cm (+/- 5-10%), 1/2 circle taper/cutting 26mm (+/- 5-10%)needle,with CE certification	No CE certificate & T8 certificate as per clause no. 5.2.12 & 5.2.9 respectively.	Not Qualified.
5	S01193 Polyester 5	Size- 5, braided, 75cm (+/- 5-10%), 1/2 circle taper/cutting 40mm (+/- 5-10%) needle,with CE certification	No CE certificate & T8 certificate as per clause no. 5.2.12 & 5.2.9 respectively.	Not Qualified.

Sl No.	Name of the Item	Specification	Reason for Rejection	Remarks
11	11	M/s. LABTECH DISPOSABLES ,Ahmedabad		
1	S02137 Sterile Blood Collection Plastic Vial	Sterile Blood Collection plastic Vials Color coded cap, clot activator, Non-vacuotainer with ISO/CE marking, capacity 5 ml fixed with a blank level displaying date manufacture and expiry, Lot No ,Manufactures details, Patient ID, Test and date. Each set	No Drug endorsement.	Not Qualified.
2	S02433 Sterile Blood Collection Plastic Vial	Sterile Blood Collection plastic Vials Color coded cap, clot activator, Non-vacuotainer with ISO/CE marking, capacity 5 ml fixed with a blank level displaying date manufacture and expiry, Lot No, Manufactures details, Patient ID, Test and date. Each set sh	No Drug endorsement.	Not Qualified.
3	S02434 Sterile Blood Collection Plastic Vial	Sterile Blood Collection plastic Vials Color coded cap, clot activator, Non-vacuotainer with ISO/CE marking, capacity 5 ml fixed with a blank level displaying date manufacture and expiry, Lot No ,Manufactures details, Patient ID, Test and date. Each set sho	No Drug endorsement.	Not Qualified.
4	S02435 Sterile Blood Collection Plastic Vial	Sterile Blood Collection plastic Vials Color coded cap, clot activator Non-vacuotainer with ISO/CE marking, capacity 5ml fixed with a blank level displaying date manufacture and expiry, LotNo, Manufactures details, Patient ID, Test and date. Each set shoul	No Drug endorsement.	Not Qualified.
5	S02436 Sterile Blood Collection Plastic Vial	Sterile Blood Collection plastic Vials Color coded cap, clot activator, Non-vacuotainer with ISO/CE marking, capacity 5 ml fixed with a blank level displaying date manufacture and expiry, Lot No ,Manufactures details, Patient ID, Test and date. Each set sh	No Drug endorsement.	Not Qualified.

Sl No.	Name of the Item	Specification	Reason for Rejection	Remarks
6	S02437 Sterile Blood Collection Plastic Vial	Sterile Blood Collection plastic VialsColor coded cap,clot activator,Non-vacuotainer with ISO/CE marking,capacity 5 ml fixed with a blank level displaying date manufacture and expiry,Lot No,Manufactures details,Patient ID,Test and date.Each set shou	No Drug endorsement.	Not Qualified.
7	S02470 Tourniquet	Each with heavy duty zipper	No Drug endorsement.	Not Qualified.

12 12 M/s. MEDIPLUS ,Odisha (Khordha)

1	S02118 Disposable Delivery Kit (ETO Sterilised)	AS PER TENDER SPECIFICATION	No T8 & No NCC of Wipro, Jackson & Medica. T8 & NCC of Adinath is older than 2 yrs. No Mfg. Lic. of Adinath.No Mfg. Lic. of Jackson for each authorized items. Mfg. Lic. of R Bandage is 4mtx10cm instead of 4mtx6cm. No BIS for Blade,Gauze & R Bandage	Not Qualified.
---	---	-----------------------------	--	----------------

13 13 M/s. TI MEDICAL PVT. LTD. ,Mumbai

1	S01126 Black Braided Silk 5-0	5-0, Curved Round Body, double armed sutures spatulated needles, with CE certification	Invalid CE certificate as per clause No. 5.2.12.	Not Qualified.
2	S01136 Polypropylene 7-0	Size-7-0, 6 / 5 mm (+/- 5-10%),length: 45cm (+/- 5-10%), 1/2 circle round bodied / Taper point, U.S.P, with CE certification	Invalid CE certificate as per clause No. 5.2.12.	Not Qualified.
3	S01138 Polypropylene 8-0	Size-8-0, 9.3 mm,length: 45cm (+/- 5-10%), 1/2 circle round bodied / Taper point, U.S.P, with CE certification	Invalid CE certificate as per clause No. 5.2.12.	Not Qualified.
4	S01183 Polyglactin 7-0	Size:- 7-0, Absorbable surgical suture USP, 0.5 metric, 30cm (+/- 5-10%), spatula, double lumen curved (A7102), with CE certification	Invalid CE certificate as per clause No. 5.2.12.	Not Qualified.
5	S01189 Polyester 1-0	Size- 1-0, 75cm (+/- 5-10%), 1/2 circle curved cutting, 36-40mm (+/- 5-10%)needle,with CE certification	Invalid CE certificate as per clause No. 5.2.12.	Not Qualified.
6	S01190 Polyester 2	Size- 2, braided, 75cm (+/- 5-10%), 1/2 circle taper/cutting 26mm (+/- 5-10%)needle,with CE certification	Invalid CE certificate as per clause No. 5.2.12.	Not Qualified.
7	S01192 Polyester 3-0	Size- 3-0, double arm needle pledgeff, 20-25mm (+/- 5-10%)needle,with CE certification	Invalid CE certificate as per clause No. 5.2.12.	Not Qualified.
8	S01193 Polyester 5	Size- 5, braided, 75cm (+/- 5-10%), 1/2 circle taper/cutting 40mm (+/- 5-10%) needle,with CE certification	Invalid CE certificate as per clause No. 5.2.12.	Not Qualified.

SI No.	Name of the Item	Specification	Reason for Rejection	Remarks
--------	------------------	---------------	----------------------	---------

14 14 M/s. JAJOO SURGICALS PVT. LTD. ,Dewas (MP)

1	S02431 Adhesive Sticker	To fix prick site to prevent bleeding, Round Shape, medicated, 22 - 25mm diameter		Qualified.
---	-------------------------	---	--	------------

Sl No.	Name of the Item	Specification	Reason for Rejection	Remarks
15	15	M/s. ROMSONS GROUP PVT. LTD. ,Delhi		
1	S02028 Corrugated Polythene Drainage Sheet	Pre-Sterile, Multichannel with radio opaque line, extra soft, non-toxic, pyrogen free, latex free, single use. Size: 25mm x250mm		Qualified.
2	S02042 Disposable PPE Kit	1. Cap. 1No. 2. Gown 1No. 3. Goggles 1No. 4. Face Mask 1No. 5. Sterile Gloves 2Pairs. 6. Shoe Cover. All the Items must be packed in a single polypack.		Qualified.
3	S02314 Flat Drain Set with trochar	7mm, 10mm. 100% silicone material, latex free; In-built non-return valve; clip for secure hanging. Radio opaque feature of drain facilitates for accurate placement; 60 cm long 100% silicone tubing for attachment to the bellow unit; capacity - 200 ml		Qualified.
4	S02331 Nebuliser set with T piece	Nebuliser-1 No., "T" Piece with 15mm Female for endotracheal tube and 22mm Female fitting for the mouth piece -1 No., flexible medical grade 2mtr long multichannel tube, kink resistant - 1 No.	No Non Conviction & T8 certificate as per clause no. 5.2.10 & 5.2.9 respectively.	Not Qualified.
5	S02343 Resusciator Kit Manual (Adult)	Ambu bag with MDI port. Star lumen kink resistant oxygen tubing, reservoir bag, pressure limiting valve and silicone mask, made up from 100% silicone, bag re-expansion property 360 degree swivel joint. Autoclavable, connectors (15/22 mm). Size - Adul		Qualified.
6	S02344 Resusciator Kit Manual (Child)	Ambu bag with MDI port. Star lumen kink resistant oxygen tubing, reservoir bag, pressure limiting valve and silicone mask, made up from 100% silicone, bag re-expansion property 360 degree swivel joint. Autoclavable, connectors (15/22 mm). Size - Chil		Qualified.
7	S02345 Resusciator Kit Manual (Neonate)	Ambu bag with MDI port. Star lumen kink resistant oxygen tubing, reservoir bag, pressure limiting valve and silicone mask, made up from 100% silicone, bag re-expansion property 360 degree swivel joint. Autoclavable, connectors (15/22 mm). Size - Infa		Qualified.
8	S02432 Bain Circuit	Sterile pack, non-toxic, pyrogen free, silicon product; sterilisable (autoclave / ETO)	Drug endorsement for material is not silicon.	Not Qualified.

Sl No.	Name of the Item	Specification	Reason for Rejection	Remarks
9	S02451 Foleys Catheter, size 8 FG	3way, Sterile pack, non-toxic, natural rubber latex, pyrogen free, size 8 FG,with CE certification.	Drug endorsement submitted for 2 way instead of 3 way & No CE certificate as per clause No. 5.2.12.	Not Qualified.
10	S02452 Foleys Catheter, size 10 FG	3way, Sterile pack, non-toxic,natural rubber latex, pyrogen free, size 10 FG,with CE certification.	Drug endorsement submitted for 2 way instead of 3 way & No CE certificate as per clause No. 5.2.12.	Not Qualified.
11	S02453 Foleys Catheter, size 12 FG	3way, Sterile pack, non-toxic,natural rubber latex, pyrogen free, size 12 FG,with CE certification.	Drug endorsement submitted for 2 way instead of 3 way & No CE certificate as per clause No. 5.2.12.	Not Qualified.
12	S02471 Urometer	2000 ml capacity, graduated, sterile, soft and kink resistant tubing, clear and transparent chamber.	No Market Standing as per clause No. 5.2.9.	Not Qualified.

16 16 M/s. SURGIPLAST (INDIA) ,Odisha (Ganjam)

1	D34023 Plaster of Paris	50Kg/Tin, plaster of paris is to be packed in a thick biodegradable polypack		Qualified.
2	S02029 Disposable Apron	Long body, sterile, disposable, Not less than 25 GSM material,Colour-Green/Blue IS 4501: 198(R2006) and ASTM F2407.	No Drug endorsement ,No (BIS & ASTM F2407), No Non Conviction & T8 certificate submitted for non sterile as per clause no. 5.2.12, 5.2.10 & 5.2.9 respectively.	Not Qualified.
3	S02055 Disposable Surgical Head Cap	Each disposable		Qualified.
4	S02079 Triple (three) layer face mask	(9cm x 17cm approx.), nose clip adaptable.(Tie on mask of no-woven, hypoallergenic 3 ply construction with filter in ply construction with filter in between offering >99% standard with Ear strings.) with IS 16289:2014.	No BIS certificate as per clause no. 5.2.12.	Not Qualified.
5	S02118 Disposable Delivery Kit (ETO Sterilised)	AS PER TENDER SPECIFICATION		Qualified.